

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McCinnis AGENCY YCSO  
DATE 6-8-11 TIME 0355  
INTOXILYZER SERIAL # 80-001222 LOCATION Verde Jail

- DM* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 10 AC.  
OR  
Pour a standard alcohol concentration solution          AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- DM* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- DM* 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- DM* 4. Air blank completed.
- DM* 5. Calibration check completed. Test results 0.089 AC.
- DM* 6. Air blank completed.
- DM* 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel McCinnis S43/QAS 33426

INTOXILYZER 8000  
Location: YCSO-VERDE JAIL  
Serial Number: 80-001222  
Core Version: 8105.48  
06/08/2011

03:55:42

-----  
Standard Lot#: 35710100A2  
Last Changed By: D. MCGINNIS #8086

-----  
QAS: D. MCGINNIS #8086  
YAVAPAI CSO

-----  
Operator: D. MCGINNIS #8086  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	03:56:14
0.100 Cal Check	0.099	03:56:35
Air Blank	0.000	03:57:04

Preventative Maintenance  
Performed -- Timer Reset