


DR 17-032055

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)  
ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SEAN CHUPP AGENCY Y. C. S. O.  
DATE 9/6/17 TIME 1806  
INTOXILYZER SERIAL # 80-001222 LOCATION CAMP VERDE JAIL

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 099 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000  
Location: YCSO - VERDE JAIL  
Serial Number: 80-001222  
Core Version: 8105.48  
09/06/2017 18:06:11

-----  
Standard Lot#: 27016100A1  
Last Changed By: J. DUBIEN #09611  
-----

DAS: S. CHUPP #9808  
YAVAPAI CSO  
-----

Operator: S. CHUPP #9808  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	18:06:43
0.100 Cal Check	0.099	18:07:04
Air Blank	0.000	18:07:33

Preventative Maintenance  
Performed -- Timer Reset

DR 17-032055

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SEAN CHUPP AGENCY Y.C.S.O.  
DATE 9/6/17 TIME 1811  
INTOXILYZER SERIAL # 80-001222 LOCATION CAMP VERDE JAIL

(X) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (X) 1. Clock time check.
- (X) 2. Date check.


OPERATIONAL TESTS

- (X) 1. Alcohol-free subject test result 0. 000 AC.
- (X) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- (X) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- (X) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

\* 550 PSI

SIGNATURE 

INTOXILYZER 8000  
Location: YCSO - VERDE JAIL  
Serial Number: 80-001222  
Core Version: 8105.48  
09/06/2017 18:17:07

---

Standard Lot#: 27016100A1  
Last Changed By: J. DUBIEN #09611

---

QAS: S. CHUPP #9808  
YAVAPAI CSO

---

Operator: S. CHUPP #9808  
YAVAPAI CSO

---

Subject: SQAP,TEST,I  
DOB: 01/01/1990  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	18:18:02
Diagnostic Test	Pass	18:18:37
Air Blank	0.000	18:19:04
0.100 Cal Check	0.099	18:19:25
Air Blank	0.000	18:19:55
Subject Test	0.025*	18:23:02
Air Blank	0.000	18:23:33
Air Blank	0.000	18:24:00
Subject Test	XXX**	18:24:19
Air Blank	0.000	18:24:52

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO - VERDE JAIL  
Serial Number: 80-001222  
Core Version: 8105.48  
09/06/2017

18:11:03

---

Standard Lot#: 27016100A1  
Last Changed By: J. DUBIEN #09611

---

QAS: S. CHUPP #9808  
YAVAPAI CSO

---

Operator: S. CHUPP #9808  
YAVAPAI CSO

---

Subject: SQAP,TEST,I  
DOB: 01/01/1990  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	18:12:14
Diagnostic Test	Pass	18:12:49
Air Blank	0.000	18:13:16
0.100 Cal Check	0.099	18:13:37
Air Blank	0.000	18:14:07
Subject Test	0.000	18:14:24
Air Blank	ABT*	18:14:33
Air Blank	RFI**	18:14:40

\*Sequence Aborted  
\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO - VERDE JAIL  
Serial Number: 80-001222  
Core Version: 8105.48  
09/06/2017 18:15:41

---

Standard Lot#: 27016100A1  
Last Changed By: J. DUBIEN #09611

---

QAS: S. CHUPP #9808  
YAVAPAI CSO

---

Operator: S. CHUPP #9808  
YAVAPAI CSO

---

Subject: SQAP,TEST,I  
DOB: 01/01/1990  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	18:16:33
Air Blank	0.000	18:17:02

\*Improper Sample


**Not a Successfully Completed Test Sequence**

22 17-03205

**EXHIBIT G-2**  
**THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**  
**ARIZONA DEPARTMENT OF PUBLIC SAFETY**  
**STANDARD QUALITY ASSURANCE PROCEDURES**  
**INTOXILYZER MODEL 8000**  
**STANDARD CALIBRATION CHECK PROCEDURE**

QA SPECIALIST SEAN CHUPP AGENCY Y.C.S.O.  
 DATE 9/6/17 TIME 1826  
 INTOXILYZER SERIAL # 80-001222 LOCATION CAMP VERDE JAIL

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
 OR  
 Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.098 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000

Location: YCSO - VERDE JAIL

Serial Number: 80-001222

Core Version: 8105.48

09/06/2017

18:26:00

-----  
Standard Lot#: 27016100A1

Last Changed By: J. DUBIEN #09611

-----  
QAS: S. CHUPP #9808

YAVAPAI CSO

-----  
Operator: S. CHUPP #9808

YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	18:26:31
0.100 Cal Check	0.098	18:26:51
Air Blank	0.000	18:27:21

Preventative Maintenance  
Performed -- Timer Reset