

146-019919

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)


ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M'QUEARY AGENCY YCSO

DATE 5/31/14 TIME _____

INTOXILYZER SERIAL # 80-001222 LOCATION CV JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 098 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

18-019919

INTOXILYZER 8000
Location: YCSO - VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
05/31/2018 20:21:13

Standard Lot#: 06717100A5
Last Changed By: J. MCQUEARY #E54

QAS: J. MCQUEARY #E54
YAVAPAI CSO

Operator: J. MCQUEARY #E54
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	20:21:44
0.100 Cal Check	0.098	20:22:05
Air Blank	0.000	20:22:35

Preventative Maintenance
Performed -- Timer Reset

146-019919

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST MCQUEARY AGENCY YCSO
DATE 5/31/18 TIME 2005
INTOXILYZER SERIAL # 80-001222 LOCATION CV JAEC

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS


- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 094 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

14-019919

INTOXILYZER 8000

Location: YCSO - VERDE JAIL

Serial Number: 80-001222

Core Version: 8105.48

05/31/2018

20:09:35

Standard Lot#: 06717100A5

Last Changed By: J. MCQUEARY #E54

QAS: J. MCQUEARY #E54

YAVAPAI CSO

Operator: J. MCQUEARY #E54

YAVAPAI CSO

Subject: SQAP, TEST

DOB: 01/01/1990

Sex: M Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	ABT*	20:10:26
Air Blank	IPS**	20:10:35

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

18-019919

INTOXILYZER 8000
Location: YCSO - VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
05/31/2018 20:12:43

Standard Lot#: 06717100A5
Last Changed By: J. MCQUEARY #E54

QAS: J. MCQUEARY #E54
YAVAPAI CSO

Operator: J. MCQUEARY #E54
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/1990
Sex: M Weight: 100

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	20:13:44
Diagnostic Test	Pass	20:14:19
Air Blank	0.000	20:14:47
0.100 Cal Check	0.098	20:15:07
Air Blank	0.000	20:15:37
Subject Test	0.123*	20:18:45
Air Blank	0.000	20:19:17
Air Blank	0.000	20:19:44
Subject Test	XXX**	20:20:06
Air Blank	RFI***	20:20:22

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

18-019919

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST McQUEARY AGENCY YCSO
DATE 5/31/18 TIME 2005
INTOXILYZER SERIAL # 80-001222 LOCATION CV JAIL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.099 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

18-019919

INTOXILYZER 8000
Location: YCSO - VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
05/31/2018 20:06:36

Standard Lot#: 06717100A5
Last Changed By: J. MCQUEARY #E54

QAS: J. MCQUEARY #E54
YAVAPAI CSO

Operator: J. MCQUEARY #E54
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	20:07:07
0.100 Cal Check	0.099	20:07:28
Air Blank	0.000	20:07:57

Preventative Maintenance
Performed -- Timer Reset

18-019919

INTOXILYZER 8000

Location: YCSO - VERDE JAIL

Serial Number: 80-001222

Core Version: 8105.48

05/31/2018

20:05:28

Standard Lot#: 06717100A5

Last Changed By: J. MCQUEARY #E54

QAS: J. MCQUEARY #E54

YAVAPAI CSO

Operator: J. MCQUEARY #E54

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	20:05:58
0.100 Cal Check	RFI*	20:06:16
Air Blank	RFI*	20:06:24

*RFI Detect