

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Chris Krogedal AGENCY YCSO
DATE 01/10/08 TIME 2130
INTOXILYZER SERIAL # 80-001455 LOCATION Bagdad

- (OK) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C.
- (OK) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (OK) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (OK) 4. Air blank completed.
- (OK) 5. Calibration check completed. Test results 0. _____ AC.
- (OK) 6. Air blank completed.
- (OK) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Chris Krogedal #570/6535

QAS# 33200

INTOXILYZER 8000

Location: YAVAPAI SO, BAGDAD

Serial Number: 80-001455

01/10/2008

21:23:33

Standard Lot#: 603901G

Last Changed By: C. KROGEDAL #6535

QAS: C. KROGEDAL #6535

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	21:24:01
Cal Check	0.100	21:24:19
Air Blank	0.000	21:24:49

C. Krogedal #570/6535

Operator's Signature

QAS #33200

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Chris Krogedal AGENCY YCSO
DATE 01/10/08 TIME 2130
INTOXILYZER SERIAL # 80-001455 LOCATION Bagdad

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. _____ AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. _____ AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

Chris Krogedal #570/6535

QAS # 33200

INTOXILYZER 8000

Location: YAVAPAI SO, BAGDAD

Serial Number : 80-001455

01/10/2008

21:26:35

Standard Lot#: 603901G

Last Changed By: C. KROGEDAL #6535

QAS: C. KROGEDAL #6535

YAVAPAI CSO

Operator: C. KROGEDAL #6535

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 01/14/1967

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:27:38
Diagnostic Test	Pass	21:28:06
Air Blank	0.000	21:28:33
0.100 Cal Check	0.100	21:28:52
Air Blank	0.000	21:29:21
Subject Test	0.000	21:30:35
Air Blank	ABT*	21:30:46
Air Blank	0.000	21:31:15

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YAVAPAI SO, BAGDAD

Serial Number : 80-001455

01/10/2008

21:32:02

Standard Lot#: 603901G

Last Changed By: C. KROGEDAL #6535

QAS: C. KROGEDAL #6535

YAVAPAI CSO

Operator: C. KROGEDAL #6535

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 01/15/1967

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	21:32:35
Air Blank	0.000	21:33:04

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YAVAPAI SO, BAGDAD

Serial Number : 80-001455

01/10/2008

21:33:25

Standard Lot#: 603901G

Last Changed By: C. KROGEDAL #6535

QAS: C. KROGEDAL #6535

YAVAPAI CSO

Operator: C. KROGEDAL #6535

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 01/15/1957

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	21:33:55
Air Blank	0.000	21:34:23

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YAVAPAI SO, BAGDAD

Serial Number : 80-001455

01/10/2008

21:35:05

Standard Lot#: 603901G

Last Changed By: C. KROGEDAL #6535

QAS: C. KROGEDAL #6535

YAVAPAI CSO

Operator: C. KROGEDAL #6535

YAVAPAI CSO

Subject: SQAP, TEST

DOB: 01/15/1967

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:35:54
Diagnostic Test	Pass	21:36:22
Air Blank	0.000	21:36:49
0.100 Cal Check	0.097	21:37:08
Air Blank	0.000	21:37:37
Subject Test	0.119*	21:40:45
Air Blank	0.000	21:41:17
Air Blank	0.000	21:41:45
Subject Test	0.039	21:42:14
Air Blank	0.000	21:42:45
Five Minute Wait		
Air Blank	0.000	21:47:41
Subject Test	XXX**	21:48:05
Air Blank	0.000	21:48:38

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence