

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Ashby #562 AGENCY YCSO  
DATE 5-26-09 TIME 0754  
INTOXILYZER SERIAL # 80-001455 LOCATION Bagdad Substation

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Deann J Ashby #562 / QAS #33400

DPS Form Exh G-2 (Rev 05-01)

INSTRUMENT NO. 8000

Location: SARAFI SQ, DUBAI

Serial Number: 80-001455

07/15/79

7500000

D. 00107 #102

8562

g/100

Time

0.006

07:54:01

0.009

07:54:20

0.000

07:54:49

*Deann J Ashby* #S62 /QAS# 33400  
Operator's Signature

EXHIBIT G-4

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST D. Ashby #502 AGENCY YCSO

DATE 5.26.09 TIME 0800

INTOXILYZER SERIAL # 80-001455 LOCATION Bagdad Substation

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

---

---

---

---

SIGNATURE Deann J Ashby #502/QAS #33400

INTOXILYZER 8000

Location: YAVAPAI SQ, BAGDAD

Serial Number: 80-001455

07:58:20

09011

D. ASHBY

1862

1862

BY

1862

Subj: 6T

DOB: 01/01/1991

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:57:10
Diagnostic Test	Pass	07:57:38
Air Blank	0.000	07:58:05
0.100 Cal Check	0.101	07:58:24
Air Blank	0.000	07:58:53
Subject Test	0.000	07:59:23
Air Blank	ABT*	07:59:36
Air Blank	0.000	07:59:59

\*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXLYZER 8000

Location: YAVAPAI SQ, BAGDAD

80-001455

08:00:03

809011

D. ASHBY

1862

1862

N

1862

SAMPLE NO: 809011-8T

DOB: 01/01/1999

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	1PS*	08:00:32
Air Blank	0.000	08:01:01

\*Improper Sample

Not a Successfully Completed Test Sequence

INTD 12/24/8000  
SCN 07 YAVAPKI SO, BAGDAD  
0-001455

08:01:22

0901T  
E. ASHBY #867  
#862

Operator: D. ASHBY #862  
YAVAPKI SO

Subject: SOAP TEST  
DOB: 01/01/1999  
Sex: M Weight: 150  
15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFT*	08:01:51
Air Blank	0.000	08:02:20

\*RFT Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: VAVAFAT CO, BAGDAD

Subject Number: 80-001455

08:00:00

7089911

D. AGRBY

4.000

4.000

B7

3.000

Sex:

Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:02:11
Diagnostic Test	Pass	08:03:24
Air Blank	0.000	08:04:11
0.100 Cal Check	0.099	08:05:16
Air Blank	0.000	08:06:00
Subject Test	0.000*	08:07:00
Air Blank	0.000	08:08:00
Air Blank	0.000	08:09:11
Subject Test	XXX**	08:09:29
Air Blank	0.000	08:10:02

\*Deficient Sample

\*\*Invalid Sample

begin new deprivation period

Not a Successfully Completed Test Sequence



003

APPAI SO, BAGLAP

00-001455

08:11:11

710901T

Py: D. ASHBY #562

#562

q/2101.

Time

0.000

08:11:41

0.095

08:12:00

0.000

08:12:29

*Deann J Ashby*  
Operator's Signature

#562 / QAS #33400