

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS
DATE 06/25/2013 TIME 0751
INTOXILYZER SERIAL # 80-001455 DPS LOCATION Bagdad Patrol

- (P/S) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (P/S) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (P/S) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (P/S) 4. Air blank completed.
- (P/S) 5. Calibration check completed. Test results 0.101 AC.
- (P/S) 6. Air blank completed.
- (P/S) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III # 6386 AZ DPS

INTOXILYZER 8000
Instrument Initialization
07:22:04 06/29/2013

INTOXILYZER 8000
Instrument Initialization
07:22:55 06/29/2013

INTOXILYZER 8000
Operator: R. BYRNE III #6386
Sample Number: 01-10-100
Date/Version: 0108.46
06/29/2013 07:51:46

Standard Lot#: 15712.0142
Last Changed By: D. ASHBY #562

QAS: R. BYRNE III #6386
AZ DPS

Operator: R. BYRNE III #6386
AZ DPS

Test	g/210L	Time
Air Blank	0.000	07:52:25
0.100 Cal Check	0.101	07:52:46
Air Blank	0.000	07:53:15

Preventative Maintenance
Performed -- Timer Reset