

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. M. Mullaney AGENCY Yavapai County Sheriff's  
DATE 9/11/13 TIME 14:25:31  
INTOXILYZER SERIAL # 80-001455 LOCATION Bigdadd 103

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution        AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(X) 4. Air blank completed.  
(X) 5. Calibration check completed. Test results 0. 101 AC.  
(X) 6. Air blank completed.  
(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Mullaney NSB/5597



INTOXILYZER 8000

Location: YAVAPAI SO, BAGDAD

Serial Number: 80-001455

Core Version: 8105.48

09/01/2013

14:25:31

-----  
Standard Lot#: 15013100A2

Last Changed By: J. MURRAY #N56  
-----

QAS: J. MURRAY #N56

YAVAPAI YCSO  
-----

Operator: J. MURRAY #N56

YAVAPAI YCSO  
-----

Test	g/210L	Time
Air Blank	0.000	14:26:03
0.100 Cal Check	0.101	14:26:24
Air Blank	0.000	14:26:54

Preventative Maintenance  
Performed -- Timer Reset



EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Murray AGENCY Yavapai County Sheriff's  
DATE 9/1/13 TIME 14:29:44  
INTOXILYZER SERIAL # 80-001455 LOCATION Bagdad 103

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS New Gas Bottle +1000 PSI  
S# - 150/3100 A2  
exp. - 6/1/2015

SIGNATURE J. Murray NS6/5597



INTOXILYZER 8000  
Location: YAVAPAI SO, BAGDAD  
Serial Number: 80-001455  
Core Version: 8105.48  
09/01/2013

14:29:44

---

Standard Lot#: 15013100A2  
Last Changed By: J. MURRAY #N56

---

QAS: J. MURRAY #N56  
YAVAPAI YCSO

---

Operator: J. MURRAY #N56  
YAVAPAI YCSO

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	14:30:48
Diagnostic Test	Pass	14:31:23
Air Blank	0.000	14:31:51
0.100 Cal Check	0.100	14:32:12
Air Blank	0.000	14:32:41
Subject Test	0.000	14:33:00
Air Blank	0.000	14:33:28
Wait	ABT*	
Air Blank	IPS**	14:33:45

\*Sequence Aborted

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**



INTOXILYZER 8000

Location: YAVAPAI SO, BAGDAD

Serial Number: 80-001455

Core Version: 8105.48

09/01/2013

14:33:48

---

Standard Lot#: 15013100A2

Last Changed By: J. MURRAY

#N56

---

QAS: J. MURRAY

#N56

YAVAPAI YCSO

---

Operator: J. MURRAY

#N56

YAVAPAI YCSO

---

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	14:34:19
Air Blank	IPS**	14:34:27

\*RFI Detect

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**



INTOXILYZER 8000  
Location: YAVAPAI SO, BAGDAD  
Serial Number: 80-001455  
Core Version: 8105.48  
09/01/2013

14:34:33

---

Standard Lot#: 15013100A2  
Last Changed By: J. MURRAY #N56

---

QAS: J. MURRAY #N56  
YAVAPAI YCSO

---

Operator: J. MURRAY #N56  
YAVAPAI YCSO

---

Subject: SOAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	14:35:47
Diagnostic Test	Pass	14:36:22
Air Blank	0.000	14:36:50
0.100 Cal Check	0.101	14:37:10
Air Blank	0.000	14:37:40
Subject Test	0.110*	14:40:48
Air Blank	0.000	14:41:20
Air Blank	0.000	14:41:47
Subject Test	XXX**	14:42:03
Air Blank	0.000	14:42:37

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**



EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Murray AGENCY Yavapai County Sheriff's  
DATE 9/11/13 TIME 14:42:52  
INTOXILYZER SERIAL # 80-001455 LOCATION Bagdad 103

- (x) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution        AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (x) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (x) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (+) 4. Air blank completed.
- (x) 5. Calibration check completed. Test results 0. 100 AC.
- (x) 6. Air blank completed.
- (x) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NS# 15597



INTOXILYZER 8000

Location: YAVAPAI SO, BAGDAD

Serial Number: 80-001455

Core Version: 8105.48

09/01/2013

14:42:52

-----  
Standard Lot#: 15013100A2

Last Changed By: J. MURRAY #N56  
-----

QAS: J. MURRAY #N56

YAVAPAI YCSO  
-----

Operator: J. MURRAY #N56

YAVAPAI YCSO  
-----

Test	g/210L	Time
Air Blank	0.000	14:43:26
0.100 Cal Check	0.100	14:43:47
Air Blank	0.000	14:44:17

Preventative Maintenance  
Performed -- Timer Reset

