

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI CO SD
DATE 1-28-18 TIME 2253
INTOXILYZER SERIAL # 80-001455 LOCATION Bagdad

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000
Location: YAUAPAI SO, BAGDAD
Serial Number: 80-001455
Core Version: 8105.48
01/28/2018

22:53:03

Standard Lot#: 27016100A1
Last Changed By: R. RIVERA #6356

QAS: R. RIVERA #6356
YAUAPAI CSO

Operator: R. RIVERA #6356
YAUAPAI CSO

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 22:53:45 |
| 0.100 Cal Check | 0.098 | 22:54:06 |
| Air Blank | 0.000 | 22:54:36 |

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI CO SO
DATE 1-28-18 TIME 2256
INTOXILYZER SERIAL # 80-001455 LOCATION BAGDAD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 093 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

INTOXILYZER 8000

Location: YAUAPAI SO, BAGDAD

Serial Number: 80-001455

Core Version: 8105.48

01/28/2018

22:56:36

Standard Lot#: 27016100A1

Last Changed By: R. RIVERA #6356

QAS: R. RIVERA #6356

YAUAPAI CSO

Operator: R. RIVERA #6356

YAUAPAI CSO

Subject: SOAP,TEST

DOB: 07/18/1980

Sex: M Weight: 123

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 22:57:35 |
| Diagnostic Test | Pass | 22:58:13 |
| Air Blank | 0.000 | 22:58:40 |
| 0.100 Cal Check | 0.097 | 22:59:01 |
| Air Blank | 0.000 | 22:59:31 |
| Subject Test | 0.000 | 23:00:04 |
| Air Blank | ABT* | 23:00:11 |
| Air Blank | IPS** | 23:00:19 |

*Sequence Aborted

**Improper Sample

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000

Location: YAUAPAI SO, BAGDAD

Serial Number: 80-001455

Core Version: 8105.48

01/28/2018

23:00:37

Standard Lot#: 27016100A1

Last Changed By: R. RIVERA #6356

QAS: R. RIVERA #6356

YAUAPAI CSO

Operator: R. RIVERA #6356

YAUAPAI CSO

Subject: SOAP, TEST

DOB: 07/18/1980

Sex: M Weight: 123

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 23:01:51 |
| Diagnostic Test | Pass | 23:02:29 |
| Air Blank | 0.000 | 23:02:56 |
| 0.100 Cal Check | 0.098 | 23:03:17 |
| Air Blank | 0.000 | 23:03:46 |
| Subject Test | 0.414* | 23:06:54 |
| Air Blank | 0.000 | 23:07:31 |
| Air Blank | 0.000 | 23:07:58 |
| Subject Test | RGE** | 23:08:28 |
| Air Blank | RFI*** | 23:08:52 |

*Deficient Sample

**Range Exceeded

***RFI Detect

Not a Successfully
Completed
Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI CO. SO
DATE 1-28-18 TIME 2309
INTOXILYZER SERIAL # 80-001455 LOCATION Basdad

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.093 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000

Location: YAUAPAI SO, BAGDAD

Serial Number: 80-001455

Core Version: 8105.48

01/28/2018

23:09:22

Standard Lot#: 27016100A1

Last Changed By: R. RIVERA #6356

QAS: R. RIVERA #6356

YAUAPAI CSO

Operator: R. RIVERA #6356

YAUAPAI CSO

Test g/210L Time

Air Blank 0.000 23:09:51
0.100 Cal Check 0.093 23:10:12
Air Blank 0.000 23:10:43

Preventative Maintenance
Performed -- Timer Reset