

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Chris Krogedal AGENCY YCSO
DATE 03/08/08 TIME 1834
INTOXILYZER SERIAL # 80-001456 LOCATION Yarnell

- (CK) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (CK) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(CK) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(CK) 4. Air blank completed.
(CK) 5. Calibration check completed. Test results 0. 101 AC.
(CK) 6. Air blank completed.
(CK) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Chris Krogedal #570/6535

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

03/08/2008

18:31:20

Standard Lot#: 730901I

Last Changed By: C. KROGEDAL #6535

QAS: C. KROGEDAL #6535

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	18:31:47
Cal Check	0.101	18:32:06
Air Blank	0.000	18:32:36


Operator's Signature

QAS # 33200