

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Ashby #562 AGENCY Yavapai County S.O.

DATE 10-26-08 TIME 0930

INTOXILYZER SERIAL # 80-001456 LOCATION Yarnell Substation

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE D. Ashby #562 / QAS # 334001

INTOXILYZER 8000  
Location: YAVAPAI SO YARNELL  
Serial Number: 80-001456  
10/26/2008 09:23:25

-----  
Standard Lot#: 730901I  
Last Changed By: C. KROGEDAL #6535  
-----

QAS: D. ASHBY #S62  
YAVAPAI CSO  
-----

| Test      | g/210L | Time     |
|-----------|--------|----------|
| Air Blank | 0.000  | 09:23:53 |
| Cal Check | 0.098  | 09:24:12 |
| Air Blank | 0.000  | 09:24:41 |

*Deann Ashby #S62 / QAS# 33400*  
Operator's Signature

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST D. Ashby #S62 AGENCY Yavapai County S.c.  
DATE 10-26-08 TIME 0932  
INTOXILYZER SERIAL # 80-001456 LOCATION Yarnell Substation

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.  
(✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.  
(✓) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(✓) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

1<sup>st</sup> attempt the set off RFI Detect did not work.  
A 2<sup>nd</sup> attempt after another test was successful -  
Received a range exceeded on the mouth alcohol  
test. A second attempt w/ less mouth wash was  
successful.

SIGNATURE Deann Ashby #S62 / QAS # 33400

INTOXILYZER 8000  
Location: YAVAPAI SO YARNELL  
Serial Number : 80-001456  
10/26/2008

09:24:49

Standard Lot#: 730901I  
Last Changed By: C. KROGEDAL #6535

QAS: D. ASHBY #S62  
YAVAPAI CSO

Operator: D. ASHBY #S62  
YAVAPAI CSO

Subject: SQAP,TEST  
DOB: 01/01/1999  
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

| Test             | g/210L | Time     |
|------------------|--------|----------|
| Air Blank        | 0.000  | 09:25:44 |
| Diagnostic Test  | Pass   | 09:26:11 |
| Air Blank        | 0.000  | 09:26:38 |
| 0.100 Cal Check  | 0.099  | 09:26:57 |
| Air Blank        | 0.000  | 09:27:27 |
| Subject Test     | 0.000  | 09:27:47 |
| Air Blank        | 0.000  | 09:28:14 |
| Five Minute Wait | ABT*   |          |
| Air Blank        | 0.000  | 09:28:53 |

\*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number : 80-001456

10/26/2008

09:28:59

Standard Lot#: 730901I

Last Changed By: C. KROGEDAL

#6535

QAS: D. ASHBY

#S62

YAVAPAI CSO

Operator: D. ASHBY

#S62

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 01/01/1999

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

| Test      | g/210L | Time     |
|-----------|--------|----------|
| Air Blank | IPS*   | 09:29:34 |
| Air Blank | 0.000  | 09:30:02 |

\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: YAVAPAI SO YARNELL  
Serial Number : 80-001456  
10/26/2008

09:30:12

Standard Lot#: 730901I  
Last Changed By: C. KROGEDAL #6535

QAS: D. ASHBY #S62  
YAVAPAI CSO

Operator: D. ASHBY #S62  
YAVAPAI CSO

Subject: SQAP,TEST  
DOB: 01/01/1999  
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 09:31:01 |
| Diagnostic Test | Pass   | 09:31:28 |
| Air Blank       | 0.000  | 09:31:55 |
| 0.100 Cal Check | 0.100  | 09:32:14 |
| Air Blank       | 0.000  | 09:32:44 |
| Subject Test    | 0.142* | 09:35:52 |
| Air Blank       | 0.000  | 09:36:25 |
| Air Blank       | 0.000  | 09:36:52 |
| Subject Test    | RGE**  | 09:37:20 |
| Air Blank       | 0.000  | 09:38:01 |

\*Deficient Sample

\*\*Range Exceeded

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number : 80-001456

10/26/2008

09:38:13

Standard Lot#: 730901I

Last Changed By: C. KROGEDAL

#6535

QAS: D. ASHBY

#S62

YAVAPAI CSO

Operator: D. ASHBY

#S62

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 01/01/1999

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

| Test      | g/210L | Time     |
|-----------|--------|----------|
| Air Blank | RFI*   | 09:39:10 |
| Air Blank | 0.000  | 09:39:38 |

\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: YAVAPAI SO YARNELL  
Serial Number : 80-001456  
10/26/2008

09:40:13

Standard Lot#: 730901I  
Last Changed By: C. KROGEDAL #6535

QAS: D. ASHBY #S62  
YAVAPAI CSO

Operator: D. ASHBY #S62  
YAVAPAI CSO

Subject: SQAP,TEST

DOB: 01/01/1999

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 09:41:03 |
| Diagnostic Test | Pass   | 09:41:30 |
| Air Blank       | 0.000  | 09:41:57 |
| 0.100 Cal Check | 0.098  | 09:42:16 |
| Air Blank       | 0.000  | 09:42:46 |
| Subject Test    | 0.054* | 09:45:54 |
| Air Blank       | 0.000  | 09:46:27 |
| Air Blank       | 0.000  | 09:46:54 |
| Subject Test    | XXX**  | 09:47:24 |
| Air Blank       | 0.000  | 09:47:58 |

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Ashby #5162 AGENCY Yavapai County S.O.  
DATE 10-26-08 TIME 0955  
INTOXILYZER SERIAL # 80-001456 LOCATION Yarnell Substation

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(✓) 4. Air blank completed.  
(✓) 5. Calibration check completed. Test results 0. 098 AC.  
(✓) 6. Air blank completed.  
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Dawn J Ashby #5162 / QAS # 33400

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

10/26/2008

09:48:16


-----  
Standard Lot#: 730901I

Last Changed By: C. KROGEDAL #6535  
-----

QAS: D. ASHBY #S62

YAVAPAI CSO  
-----

| Test      | g/210L | Time     |
|-----------|--------|----------|
| Air Blank | 0.000  | 09:48:43 |
| Cal Check | 0.098  | 09:49:02 |
| Air Blank | 0.000  | 09:49:31 |

 #SL62/QAS# 33400  
Operator's Signature