

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Ashby #562 AGENCY YCSO
DATE 1-25-09 TIME 1341
INTOXILYZER SERIAL # 30-001456 LOCATION Yarnell Substation

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Deann J Ashby #562

DPS Form Exh G-2 (Rev 05-01)

INSTRUMENT ID: _____
LOCATION: YARNELL SUBSTATION
SERIAL NUMBER: 30-001456
01/25/2009 13:41:32
Standard Lot#: 7309011
Last Changed By: C. KRUGERDAL #5535
DPS: D. ASHBY #562
YCSO

Test	g/21.0L	Time
Air Blank	0.000	13:41:59
Cal Check	0.100	13:42:18
Air Blank	0.000	13:42:46

Deann J Ashby
Operator's Signature
#562 QAS#3340

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST D. Ashby #5102 AGENCY YCSO

DATE 1-25-09 TIME 1353

INTOXILYZER SERIAL # 80-001456 LOCATION Yarnell Substation

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Dream A Ashby #5102 QAS# 33400

INTOXILYZER 8000
Location: YAUAPAI SO YARNELL
Serial Number: 80-001456
01/25/2009 13:46:30

Standard Lot#: 7309011
Last Changed By: C. KROGEDAL #6535

OAS: D. ASHBY #562
YAUAPAI CSO

Operator: D. ASHBY #562
YAUAPAI CSO

Subject: SOAP, TEST
DOB: 01/01/1999
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:47:23
Diagnostic Test	Pass	13:47:52
Air Blank	0.000	13:48:19
0.100 Cal Check	0.099	13:48:38
Air Blank	0.000	13:49:06
Subject Test	0.000	13:49:25
Air Blank	0.000	13:49:53
Five Minute Wait	ABT*	
Air Blank	0.000	13:50:27

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: YAUAPAI SO YARNELL
Serial Number: 80-001456
01/25/2009 13:54:13

Standard Lot#: 7309011
Last Changed By: C. KROGEDAL #6535

OAS: D. ASHBY #562
YAUAPAI CSO

Operator: D. ASHBY #562
YAUAPAI CSO

Subject: SOAP, TEST
DOB: 01/01/1999
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:55:11
Diagnostic Test	Pass	13:55:39
Air Blank	0.000	13:56:07
0.100 Cal Check	0.099	13:56:26
Air Blank	0.000	13:56:55
Subject Test	0.200*	14:00:04
Air Blank	0.000	14:00:37
Air Blank	0.000	14:01:04
Subject Test	XXX**	14:01:25
Air Blank	0.000	14:01:59

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: YAUAPAI SO YARNELL
Serial Number: 80-001456
01/25/2009 13:51:13

Standard Lot#: 7309011
Last Changed By: C. KROGEDAL #6535

OAS: D. ASHBY #562
YAUAPAI CSO

Operator: D. ASHBY #562
YAUAPAI CSO

Subject: SOAP, TEST
DOB: 01/01/1999
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	13:52:06
Air Blank	0.000	13:52:34

*Improper Sample

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: YAUAPAI SO YARNELL
Serial Number: 80-001456
01/25/2009 13:52:53

Standard Lot#: 7309011
Last Changed By: C. KROGEDAL #6535

OAS: D. ASHBY #562
YAUAPAI CSO

Operator: D. ASHBY #562
YAUAPAI CSO

Subject: SOAP, TEST
DOB: 01/01/1999
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	13:53:32
Air Blank	0.000	13:54:00

RFI Detect

Not a Successfully
Completed
Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Ashby #562 AGENCY YCSO
 DATE 1-25-09 TIME 1403
 INTOXILYZER SERIAL # 80-001456 LOCATION Yarnell Substation

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
 OR
 Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 0.98 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Deann J Ashby #562 QAS#33400

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
 Location: YUAPPAI SO YARNELL
 Serial Number: 80-001456
 9/125/2009 14:02:35

Standard Lot#: 7309011
 Last Changed By: C. KRUGERPL #6535

DPS: D. ASHBY #562
 YUAPPAI CSO

Test	9/2/08	Time
Air Blank	0.000	14:03:02
Cal Check	0.998	14:03:21
Air Blank	0.000	14:03:51

Deann J Ashby
 Operator's Signature
 #562 QAS#33400