

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. V. [unclear] AGENCY Yavapai County Sheriff's
DATE 9/11/73 TIME 16:04:49
INTOXILYZER SERIAL # 80-001456 LOCATION Yarnell 103

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution of 100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 096 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. M. [unclear] NS6/5597

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

09/01/2013

16:04:49

Standard Lot#: 13712100A2

Last Changed By: D. ASHBY #S62

QAS: J. MURRAY #N56

YAVAPAI YCSO

Operator: J. MURRAY #N56

YAVAPAI YCSO

Test g/210L Time

Air Blank 0.000 16:05:13
0.100 Cal Check 0.096 16:05:19
Air Blank 0.000 16:06:14

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Mussad AGENCY Yavapai County Sheriff's
DATE 9/1/13 TIME 15:08:49
INTOXILYZER SERIAL # 80-001456 LOCATION Idson 103

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

13712100AZ
exp 7/1/2014
-900 PST.

SIGNATURE J. Mussad NSB/5597

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

09/01/2013

16:08:49

Standard Lot#: 13712100A2

Last Changed By: D. ASHBY

#362

QAS: J. MURRAY

#N56

YAVAPAI YCSO

Operator: J. MURRAY

#N56

YAVAPAI YCSO

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:09:42
Diagnostic Test	Pass	16:10:17
Air Blank	0.000	16:10:40
0.100 Cal Check	0.095	16:11:05
Air Blank	0.000	16:11:35
Subject Test	0.000	16:12:08
Air Blank	0.000	16:12:36
Wait	ABT*	
Air Blank	IPS**	16:12:55

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

09/01/2013

16:12:58

Standard Lot#: 13712100A2

Last Changed By: D. ASHBY

#562

QAS: J. MURRAY

#N56

YAVAPAI YCSO

Operator: J. MURRAY

#N56

YAVAPAIYCSO 5597

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	16:13:58
Air Blank	IPS**	16:14:05

*RFI Detect

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YAVAPAI SO YARNELL
Serial Number: 80-001456
Core Version: 8105.48
09/01/2013

16:14:44

Standard Lot#: 13712100A2
Last Changed By: D. ASHBY #SG2

QAS: J. MURRAY #N56
YAVAPAI YCSO

Operator: J. MURRAY #5597
YAVAPAI YCSO56

Subject: SQAP,TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:15:44
Diagnostic Test	Pass	16:16:19
Air Blank	0.000	16:16:47
0.100 Cal Check	0.096	16:17:07
Air Blank	0.000	16:17:37
Subject Test	0.108*	16:20:44
Air Blank	0.000	16:21:16
Air Blank	0.000	16:21:44
Subject Test	XXX**	16:22:10
Air Blank	0.000	16:23:13

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence



EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. V. Slay AGENCY Yavapai County Sheriff's
DATE 9/11/13 TIME 16:23:28
INTOXILYZER SERIAL # 80-001456 LOCATION Yainell 103

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0.096 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. M. V. Slay NSB/5597

INTOXILYSER 8000

Location: YAVAPAI SO YARNEIL

Serial Number: 80-001456

Core Version: 8105.48

09/01/2013

16:23:23

Standard Lot#: 13712100A2

Last Changed By: D. ASHBY #862

QAS: J. MURRAY #5597

YAVAPAI YCSON56

Operator: J. MURRAY #5597

YAVAPAI YCSON56

Test g/210L Time

Air Blank 0.000 16:23:56
0.100 Cal Check 0.096 16:24:00
Air Blank 0.000 16:24:00

Preventative Maintenance
Performed -- Timer Reset

