

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI County Sheriff's  
DATE 6-15-16 TIME 0002  
INTOXILYZER SERIAL # 80-001456 LOCATION Yarwell

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
 4. Air blank completed.  
 5. Calibration check completed. Test results 0.097 AC.  
 6. Air blank completed.  
 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 6356

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

06/15/2016

00:02:25

-----  
Standard Lot#: 36314100A4

Last Changed By: D. ASHBY #2271

-----  
QAS: R. RIVERA #6356

YAVAPAI CSO

-----  
Operator: R. RIVERA #6356

YAVAPAI CSO

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      00:02:56  
0.100 Cal Check                0.097                      00:03:16  
Air Blank                      0.000                      00:03:46

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI COUNTY SHERIFFS  
DATE 6-15-16 TIME 0008  
INTOXILYZER SERIAL # 80-001456 LOCATION YARVELL

(  T. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  1. Clock time check.  
(  2. Date check.

OPERATIONAL TESTS

- (  1. Alcohol-free subject test result 0. 000 AC.  
(  2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(  3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(  4. Standard Calibration Check standard 0. 100 AC: Result 0 095 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

---

---

---

---

SIGNATURE 

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

06/15/2016

00:08:51

---

Standard Lot#: 36314100A4

Last Changed By: D. ASHBY

#2271

---

QAS: R. RIVERA

#6356

YAVAPAI CSO

---

Operator: R. RIVERA

#6356

YAVAPAI CSO

---

Subject: SQAP,TEST

DOB: 07/18/1980

Sex: M

Weight: 123

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	00:10:00
Diagnostic Test	Pass	00:10:36
Air Blank	0.000	00:11:03
0.100 Cal Check	0.096	00:11:24
Air Blank	0.000	00:11:54
Subject Test	0.000	00:12:19
Air Blank	0.000	00:12:47
Wait	ABT*	
Air Blank	IPS**	00:13:07

\*Sequence Aborted

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

06/15/2016

00:15:00

---

Standard Lot#: 36314100A4

Last Changed By: D. ASHBY

#2271

---

OAS: R. RIVERA

#6356

YAVAPAI CSO

---

Operator: R. RIVERA

#6356

YAVAPAI CSO

---

Subject: SQAP,TEST

DOB: 07/18/1980

Sex: M

Weight: 123

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	00:16:11
Diagnostic Test	Pass	00:16:46
Air Blank	0.000	00:17:14
0.100 Cal Check	0.097	00:17:34
Air Blank	0.000	00:18:05
Subject Test	0.392*	00:21:12
Air Blank	0.000	00:21:54
Air Blank	0.000	00:22:21
Subject Test	RGE**	00:22:47
Air Blank	RFI***	00:22:58

\*Deficient Sample

\*\*Range Exceeded

\*\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI COUNTY SHERIFF'S OFFICE  
DATE 6-15-16 TIME 0025  
INTOXILYZER SERIAL # 80-001456 LOCATION YARWELL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.095 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 





INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

06/15/2016

00:31:59

-----  
Standard Lot#: 36314100A4

Last Changed By: D. ASHBY #2271

-----  
QAS: R. RIVERA #6356

YAVAPAI CSO

-----  
Operator: R. RIVERA #6356

YAVAPAI CSO

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      00:34:36  
0.100 Cal Check                0.095                      00:34:56  
Air Blank                      0.000                      00:35:26

Preventative Maintenance  
Performed -- Timer Reset