

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI CO. SO  
DATE 9-11-16 TIME 2055  
INTOXILYZER SERIAL # 80-001456 LOCATION YARNELL

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 095 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 6356



EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. Ruena AGENCY Yavapai County S.O.  
DATE 9-11-16 TIME 2:23  
INTOXILYZER SERIAL # 80-001454 LOCATION YAVAPAI

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.095 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 90-001456

Core Version: 8105.48

09/11/2016

21:23:12

-----  
Standard Lot#: 36314100A4

Last Changed By: D. ASHBY #2271

-----  
QAS: R. RIVERA #6356

YAVAPAI CSO

-----  
Operator: R. RIVERA #6356

YAVAPAI CSO

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      21:23:43  
0.100 Cal Check                0.095                      21:23:04  
Air Blank                      0.000                      21:24:30

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI COUNTY S.O.  
DATE 9-11-16 TIME 2100  
INTOXILYZER SERIAL # 80-00456 LOCATION YARNELL

( 1. Display Reads "PUSH BUTTON TO START")

DIAGNOSTIC TESTS

- ( 1. Clock time check.
- ( 2. Date check.

OPERATIONAL TESTS

- ( 1. Alcohol-free subject test result 0. 000 AC.
- ( 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- ( 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- ( 4. Standard Calibration Check standard 0. 100 AC. Result 0. 095 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE [Signature]

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

09/11/2016

21:00:11

---

Standard Lot#: 36314100A4

Last Changed By: D. ASHBY #2271

---

QAS: R. RIVERA #6356

YAVAPAI CSO

---

Operator: R. RIVERA #6356

YAVAPAI CSO

---

Subject: SQAP,TEST

DOB: 07/18/1980

Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	21:01:33
Diagnostic Test	Pass	21:02:08
Air Blank	0.000	21:02:36
0.100 Cal Check	0.095	21:02:57
Air Blank	0.000	21:03:26
Subject Test	0.000	21:04:09
Air Blank	0.000	21:04:36
Wait	ABT*	
Air Blank	IPS**	21:04:51

\*Sequence Aborted

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YAVAPAI SO YARNELL  
Serial Number: 80-001456  
Core Version: 8105.48  
09/11/2016

21:06:09

---

Standard Lot#: 36314100A4  
Last Changed By: D. ASHBY #2271

---

OAS: R. RIVERA #6356  
YAVAPAI CSO

---

Operator: R. RIVERA #6356  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 07/18/1980  
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	21:08:10
Diagnostic Test	Pass	21:08:45
Air Blank	0.000	21:09:12
0.100 Cal Check	0.096	21:09:33
Air Blank	0.000	21:10:03
Subject Test	0.081*	21:13:10
Air Blank	0.000	21:13:48
Air Blank	0.000	21:14:16
Subject Test	XXX**	21:14:54
Air Blank	0.000	21:15:32

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YAVAPAI SO YARNELL  
Serial Number: 80-001456  
Core Version: 8105.48  
09/11/2016

21:16:16

---

Standard Lot#: 36314100A4  
Last Changed By: D. ASHBY #2271

---

QAS: R. RIVERA #6356  
YAVAPAI CSO

---

Operator: R. RIVERA #6356  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 07/18/1980  
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	21:17:15
Diagnostic Test	Pass	21:17:50
Air Blank	0.000	21:18:17
0.100 Cal Check	0.095	21:18:38
Air Blank	0.000	21:19:08
Subject Test	0.009	21:20:11
Air Blank	RFI*	21:20:36
Air Blank	RFI*	21:20:52

\*RFI Detect

Not a Successfully Completed Test Sequence