

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. RIVERA AGENCY YAVAPAI CO. SO
DATE 7-28-17 TIME 1546
INTOXILYZER SERIAL # 80-001456 LOCATION YARNELL

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 098 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

17-027141

INTOXILYZER 8000
Location: YAUAPAI SO YARNELL
Serial Number: 80-001456
Core Version: 8105.48
07/28/2017

15:46:47

Standard Lot#: 27016100A1
Last Changed By: D. ASHBY #2271

DAS: R. RIVERA #6356
YAUAPAI CSO

Operator: R. RIVERA #6356
YAUAPAI CSO

Test	g/210L	Time
Air Blank	0.000	15:47:19
0.100 Cal Check	0.098	15:47:40
Air Blank	0.000	15:48:10

Preventative Maintenance
Performed -- Timer Reset