

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. RIVERA AGENCY YAVAPAI CO. SO  
DATE 9-13-17 TIME 1803  
INTOXILYZER SERIAL # 80-001456 LOCATION YARNELL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 5000  
Location: YAUAPAI SO YARNELL  
Serial Number: 80-001456  
Core Version: 8195.48  
09/13/2017

18:03:50

-----  
Standard Lot#: 27016100A1  
Last Changed By: D. ASHBY #2271  
-----

QAS: R. RIVERA #6356  
YAUAPAI CSO  
-----

Operator: R. RIVERA #6356  
YAUAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	18:04:21
0.100 Cal Check	0.098	18:04:42
Air Blank	0.000	18:05:12

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI CO. SO  
DATE 9-13-17 TIME 1805  
INTOXILYZER SERIAL # 80-001456 LOCATION YARNEN

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

SQAP conducted to put back into service due to  
power outage

SIGNATURE 

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

09/13/2017

18:05:48

-----  
Standard Lot#: 27016100A1

Last Changed By: D. ASHBY #2271  
-----

QAS: R. RIVERA #6356

YAVAPAI CSO  
-----

Operator: R. RIVERA #6356

YAVAPAI CSO  
-----

Subject: SQAP,TEST

DOB: 07/18/1980

Sex: M Weight: 123  
-----

15 Minute Deprivation Period? Yes  
-----

Test	g/210L	Time
Air Blank	0.000	18:06:47
Diagnostic Test	Pass	18:07:25
Air Blank	0.000	18:07:52
0.100 Cal Check	0.097	18:08:13
Air Blank	0.000	18:08:43
Subject Test	0.000	18:09:03
Air Blank	ABT*	18:09:10
Air Blank	IPS**	18:09:18

\*Sequence Aborted

\*\*Improper Sample

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: YAUAPAI SO YARNELL  
Serial Number: 80-001456  
Core Version: 8105.48  
09/13/2017

18:24:13

-----  
Standard Lot#: 27016100A1  
Last Changed By: D. ASHBY #2271  
-----

QAS: R. RIVERA #6356  
YAUAPAI CSO  
-----

Operator: R. RIVERA #6356  
YAUAPAI CSO  
-----

Subject: SQAP, TEST  
DOB: 07/18/1980  
Sex: M      Weight: 123  
-----

15 Minute Deprivation Period? Yes  
-----

Test	g/210L	Time
Air Blank	0.000	18:25:38
Diagnostic Test	Pass	18:26:15
Air Blank	0.000	18:26:42
0.100 Cal Check	0.096	18:27:03
Air Blank	0.000	18:27:33
Subject Test	0.133*	18:30:41
Air Blank	0.000	18:31:19
Air Blank	0.000	18:31:46
Subject Test	RGE**	18:32:07
Air Blank	RFI***	18:32:19

\*Deficient Sample  
\*\*Range Exceeded  
\*\*\*RFI Detect

Not a Successfully  
Completed

Test Sequence

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI CO. SO  
DATE 9-13-17 TIME 1846  
INTOXILYZER SERIAL # 80-001456 LOCATION YAVAPAI

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.097 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: YAUAPAI SO YARNELL  
Serial Number: 80-001456  
Core Version: 8105.48  
09/13/2017

18:46:26

-----  
Standard Lot#: 27016100A1  
Last Changed By: D. ASHBY #2271  
-----

QAS: R. RIVERA #6356  
YAUAPAI CSO  
-----

Operator: R. RIVERA #6356  
YAUAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	18:46:56
0.100 Cal Check	0.097	18:47:17
Air Blank	0.000	18:47:47

Preventative Maintenance  
Performed -- Timer Reset