

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. Rivera 6356 AGENCY YAVAPAI CO SO
DATE 12-27-17 TIME 1506
INTOXILYZER SERIAL # 80-001456 LOCATION YARNUM

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000
Location: YAUAPAI SO YARNELL
Serial Number: 80-001456
Core Version: 8105.48
12/27/2017 15:06:41

Standard Lot#: 27016100A1
Last Changed By: D. ASHBY #2271

QAS: R. RIVERA #6356
YAUAPAI CSO

Operator: R. RIVERA #6356
YAUAPAI CSO

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 15:07:09 |
| 0.100 Cal Check | 0.097 | 15:07:30 |
| Air Blank | 0.000 | 15:08:00 |

Preventative Maintenance
Performed -- Timer Reset