

18-029327

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R Rivera 6356 AGENCY YAVAPAI CO SO
DATE 8-8-18 TIME 2226
INTOXILYZER SERIAL # 80-001456 LOCATION YARNELL

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.095 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  6356

18-029327

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

08/08/2018

22:26:50

Standard Lot#: 27016100A1

Last Changed By: D. ASHBY #2271

QAS: R. RIVERA #6356

YAVAPAI CSO

Operator: R. RIVERA #6356

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	22:29:16
0.100 Cal Check	0.095	22:29:37
Air Blank	0.000	22:30:06

Preventative Maintenance
Performed -- Timer Reset

18-029327

INTOXILYZER 8000
Location: YAVAPAI SO YARNELL
Serial Number: 80-001456
Core Version: 8105.48
08/08/2018 22:35:12

Standard Lot#: 27016100A1
Last Changed By: D. ASHBY #2271

QAS: R. RIVERA #6356
YAVAPAI CSO

Operator: R. RIVERA #06356
YCSO

Subject: SQAP, TESH
DOB: 07/18/1980
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:36:42
Diagnostic Test	Pass	22:37:17
Air Blank	0.000	22:37:44
0.100 Cal Check	0.095	22:38:05
Air Blank	0.000	22:38:35
Subject Test	0.000	22:38:55
Air Blank	ABT*	22:39:01
Air Blank	IPS**	22:39:09

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

18-029327

INTOXILYZER 8000
Location: YAVAPAI SO YARNELL
Serial Number: 80-001456
Core Version: 8105.48
08/08/2018 22:39:16

Standard Lot#: 27016100A1
Last Changed By: D. ASHBY #2271

QAS: R. RIVERA #6356
YAVAPAI CSO

Operator: R. RIVERA #6356
YAVAPAI CSO

Subject: SQAP, TEST
DOB: 07/18/1980
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:41:06
Diagnostic Test	Pass	22:41:41
Air Blank	0.000	22:42:08
0.100 Cal Check	0.095	22:42:29
Air Blank	0.000	22:42:59
Subject Test	0.051*	22:46:06
Air Blank	0.000	22:46:45
Air Blank	0.000	22:47:13
Subject Test	XXX**	22:47:46
Air Blank	RFI***	22:48:05

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

18-029327

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST R. Rivera AGENCY YAVAPAI CO SO
DATE 8-8-18 TIME 2235
INTOXILYZER SERIAL # 80-001456 LOCATION YARNELL

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 093 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE  6354

18-029327

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. Rivera 6356 AGENCY YAVAPAI CO SO

DATE 8-8-18 TIME 2248

INTOXILYZER SERIAL # 80-001456 LOCATION YARNEN

() 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

() 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

() 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

() 4. Air blank completed.

() 5. Calibration check completed. Test results 0.093 AC.

() 6. Air blank completed.

() 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

18-029327

INTOXILYZER 8000

Location: YAVAPAI SO YARNELL

Serial Number: 80-001456

Core Version: 8105.48

08/08/2018

22:48:35

Standard Lot#: 27016100A1

Last Changed By: D. ASHBY #2271

QAS: R. RIVERA #6356

YAVAPAI CSO

Operator: R. RIVERA #6356

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	22:49:05
0.100 Cal Check	0.093	22:49:26
Air Blank	0.000	22:49:56

Preventative Maintenance
Performed -- Timer Reset