

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST MURPHY NSB AGENCY Y250  
DATE 9-16-08 TIME 9 22 14  
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.098 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murphy NSB

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

09/16/2008

09:22:14

-----  
Standard Lot#: 735401G

Last Changed By: J. MURRAY #N-56  
-----

QAS: J. MURRAY #N-56

YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	09:22:41
Cal Check	0.098	09:23:00
Air Blank	0.000	09:23:29

*J. MURRAY N56*

-----  
Operator's Signature

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. MURRAY NSB AGENCY YESO  
DATE 9-16-08 TIME 92548  
INTOXILYZER SERIAL # 80-01234 LOCATION Seligman

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS Gas Standard - 7354016  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE J. MURRAY NSB

INTOXILYZER 8000  
Location: YCSO SELIGMAN  
Serial Number : 80-001734  
09/16/2008

09:25:48

Standard Lot#: 735401G  
Last Changed By: J. MURRAY #N-56

QAS: J. MURRAY #N-56  
YAVAPAI CSO

Operator: J. MURRAY #N-56  
YAVAPAI CSO

Subject: SQAP,TEST  
DOB: 11/20/1961  
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:27:11
Diagnostic Test	Pass	09:27:39
Air Blank	RFI*	09:27:54
Air Blank	IPS**	09:28:09

\*RFI Detect  
\*\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: YCSO SELIGMAN  
Serial Number : 80-001734  
09/16/2008

09:28:34

Standard Lot#: 735401G  
Last Changed By: J. MURRAY #N-56

QAS: J. MURRAY #N-56  
YAVAPAI CSO

Operator: J. MURRAY #N-56  
YAVAPAI CSO

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	ABT*	09:29:27
Air Blank	RFI**	09:29:40

\*Sequence Aborted  
\*\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: YCSO SELIGMAN  
Serial Number : 80-001734  
09/16/2008

09:30:16

Standard Lot#: 735401G  
Last Changed By: J. MURRAY #N-56

QAS: J. MURRAY #N-56  
YAVAPAI CSO

Operator: J. MURRAY #N-56  
YAVAPAI CSO

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:31:08
Diagnostic Test	Pass	09:31:36
Air Blank	0.000	09:32:03
0.100 Cal Check	0.099	09:32:22
Air Blank	0.000	09:32:51
Subject Test	0.000	09:33:07
Air Blank	0.000	09:33:34
Five Minute Wait		
Air Blank	0.000	09:38:35
Subject Test	0.286*	09:41:43
Air Blank	0.000	09:42:18

\*Deficient Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO SELIGMAN  
Serial Number : 80-001734  
09/16/2008 09:42:54

Standard Lot#: 735401G  
Last Changed By: J. MURRAY #N-56

QAS: J. MURRAY #N-56  
YAVAPAI CSO

Operator: J. MURRAY #N-56  
YAVAPAI CSO

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:43:53
Diagnostic Test	Pass	09:44:21
Air Blank	0.000	09:44:48
0.100 Cal Check	0.097	09:45:07
Air Blank	0.000	09:45:37
Subject Test	XXX*	09:46:17
Air Blank	0.000	09:46:53

\*Invalid Sample  
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST MURRAY NSB AGENCY YCSO  
DATE 9-16-08 TIME 958 24  
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(X) 4. Air blank completed.  
(X) 5. Calibration check completed. Test results 0.098 AC.  
(X) 6. Air blank completed.  
(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. MURRAY NSB

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

09/16/2008

09:58:24

-----  
Standard Lot#: 735401G

Last Changed By: J. MURRAY #N-56  
-----

QAS: J. MURRAY #N-56

YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	09:58:52
Cal Check	0.098	09:59:10
Air Blank	0.000	09:59:40

J. Murray NS6

Operator's Signature