

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. MURPHY AGENCY YCSO
DATE 3-28-10 TIME 93643
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

- (x) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (x) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (x) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (x) 4. Air blank completed.
- (x) 5. Calibration check completed. Test results 0.099 AC.
- (x) 6. Air blank completed.
- (x) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. MURPHY NSB

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

03/28/2010

09:36:43

Standard Lot#: 901401I

Last Changed By: J. MURRAY #5597

QAS: J. MURRAY #5597

YAVAPAI CSO

Operator: J. MURRAY #5597

YAVAPAI CSO

Test g/210L Time

Air Blank 0.000 09:37:21
0.100 Cal Check 0.099 09:37:41
Air Blank 0.000 09:38:11

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. Murray AGENCY YCSO
DATE 3-28-10 TIME 9:57:05
INTOXILYZER SERIAL # 80-01734 LOCATION Saligman

(x) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (x) 1. Clock time check.
- (x) 2. Date check.

OPERATIONAL TESTS

- (x) 1. Alcohol-free subject test result 0. 000 AC.
- (x) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (x) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (x) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES X NO

COMMENTS

901401T -500 PSI

SIGNATURE J. Murray

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

03/28/2010

09:57:05

Standard Lot#: 901401I

Last Changed By: J. MURRAY

#5597

OAS: J. MURRAY

#5597

YAVAPAI CSO

Operator: J. MURRAY

#5597

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:58:48
Diagnostic Test	Pass	09:59:23
Air Blank	0.000	09:59:50
0.100 Cal Check	0.099	10:00:11
Air Blank	0.000	10:00:40
Subject Test	0.000	10:01:14
Air Blank	0.000	10:01:42
Wait	ABT*	
Air Blank	RFI**	10:04:26

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO SELIGMAN
Serial Number: 80-001734
Core Version: 8105.48
03/28/2010 10:05:37

Standard Lot#: 901401I
Last Changed By: J. MURRAY #5597

QAS: J. MURRAY #5597
YAVAPAI CSO

Operator: J. MURRAY #5597
YAVAPAI CSO

Subject: TEST, SQAP
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	10:06:10
Air Blank	RFI**	10:06:21

*Improper Sample
**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO SELIGMAN
Serial Number: 80-001734
Core Version: 8105.48
03/28/2010

10:06:54

Standard Lot#: 901401I
Last Changed By: J. MURRAY #5597

QAS: J. MURRAY #5597
YAVAPAI CSO

Operator: J. MURRAY #5597
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:07:49
Diagnostic Test	Pass	10:08:24
Air Blank	0.000	10:08:52
0.100 Cal Check	0.099	10:09:13
Air Blank	0.000	10:09:42
Subject Test	0.075*	10:12:49
Air Blank	0.000	10:13:25
Air Blank	0.000	10:13:52
Subject Test	XXX**	10:14:30
Air Blank	0.000	10:15:06

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO SELIGMAN
Serial Number: 80-001734
Core Version: 8105.48
03/28/2010

10:17:46

Standard Lot#: 901401I
Last Changed By: J. MURRAY #5597

QAS: J. MURRAY #5597
YAVAPAI CSO

Operator: J. MURRAY #5597
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:18:40
Diagnostic Test	Pass	10:19:15
Air Blank	0.000	10:19:43
0.100 Cal Check	0.098	10:20:03
Air Blank	0.000	10:20:33
Subject Test	XXX*	10:20:58
Air Blank	0.000	10:21:37

*Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

03/28/2010

10:21:46

Standard Lot#: 901401I

Last Changed By: J. MURRAY

#5597

QAS: J. MURRAY

#5597

YAVAPAI CSO

Operator: J. MURRAY

#5597

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:22:38
Diagnostic Test	Pass	10:23:13
Air Blank	0.000	10:23:40
0.100 Cal Check	0.098	10:24:01
Air Blank	0.000	10:24:31
Subject Test	RGE*	10:25:07
Air Blank	0.000	10:25:45

*Range Exceeded

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray AGENCY YCSO
DATE 3-28-10 TIME 10 26 18
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 098 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

03/28/2010

10:26:18

Standard Lot#: 901401I

Last Changed By: J. MURRAY #5597

QAS: J. MURRAY #5597

YAVAPAI CSO

Operator: J. MURRAY #5597

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	10:26:51
0.100 Cal Check	0.098	10:27:12
Air Blank	0.000	10:27:42

Preventative Maintenance
Performed -- Timer Reset