

COPY

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray NS6 AGENCY Yavapai County Sheriff's  
DATE 2/9/11 TIME 11:00:30  
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution        AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.099 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NS6

INTOXILYZER 8000  
Location: YCSO SELIGMAN  
Serial Number: 80-001734  
Core Version: 8105.48  
02/09/2011

11:00:30

-----  
Standard Lot#: 779386  
Last Changed By: J. MURRAY #N56  
-----

QAS: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Operator: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Test	g/210L	Time
Air Blank	0.000	11:01:50
0.100 Cal Check	0.099	11:02:11
Air Blank	0.000	11:02:40

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. Murray NSB AGENCY Yavapai County Sheriff's  
DATE 2/9/11 TIME 11:13:02  
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

Gas Tank - 1000 PSI

SIGNATURE J. Murray NSB

INTOXILYZER 8000  
Location: YCSO SELIGMAN  
Serial Number: 80-001734  
Core Version: 8105.48  
02/09/2011

11:13:02

---

Standard Lot#: 779386  
Last Changed By: J. MURRAY #N56

---

QAS: J. MURRAY #N56  
YAVAPAI YCSO

---

Operator: J. MURRAY #N56  
YAVAPAI YCSO

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	11:14:04
Diagnostic Test	Pass	11:14:39
Air Blank	0.000	11:15:06
0.100 Cal Check	0.098	11:15:27
Air Blank	0.000	11:15:57
Subject Test	0.000	11:17:09
Air Blank	0.000	11:17:36
Wait	ABT*	
Air Blank	RFI**	11:18:15

\*Sequence Aborted  
\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO SELIGMAN  
Serial Number: 80-001734  
Core Version: 8105.48  
02/09/2011

11:18:27

---

Standard Lot#: 779386  
Last Changed By: J. MURRAY #N56

---

QAS: J. MURRAY #N56  
YAVAPAI YCSO

---

Operator: J. MURRAY #N56  
YAVAPAI YCSO

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	ABT*	11:19:05
Air Blank	RFI**	11:19:16

\*Sequence Aborted

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO SELIGMAN  
Serial Number: 80-001734  
Core Version: 8105.48  
02/09/2011

11:19:53

---

Standard Lot#: 779386  
Last Changed By: J. MURRAY #N56

---

QAS: J. MURRAY #N56  
YAVAPAI YCSO

---

Operator: J. MURRAY #N56  
YAVAPAI YCSO

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	11:20:49
Diagnostic Test	Pass	11:21:24
Air Blank	0.000	11:21:52
0.100 Cal Check	0.099	11:22:12
Air Blank	0.000	11:22:42
Subject Test	0.191*	11:25:49
Air Blank	0.000	11:26:24
Air Blank	0.000	11:26:51
Subject Test	XXX**	11:27:26
Air Blank	0.000	11:28:03

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

02/09/2011

11:30:29

---

Standard Lot#: 779386

Last Changed By: J. MURRAY

#N56

---

QAS: J. MURRAY

#N56

YAVAPAI YCSO

---

Operator: J. MURRAY

#N56

YAVAPAI YCSO

---

Subject: SQAP,TEAST

DOB: 10/20/1960

Sex: F

Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	11:31:11
Air Blank	IPS*	11:31:22

\*Improper Sample

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. MURPHY NSB AGENCY Yavapai County Sheriff's  
DATE 2/9/11 TIME 11:32:19  
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.  
Pour a standard alcohol concentration solution OR X AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. MURPHY NSB

INTOXILYZER 8000  
Location: YCSO SELIGMAN  
Serial Number: 80-001734  
Core Version: 8105.48  
02/09/2011

11:32:19

-----  
Standard Lot#: 779386  
Last Changed By: J. MURRAY #N56  
-----

QAS: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Operator: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Test	g/210L	Time
Air Blank	0.000	11:32:55
0.100 Cal Check	0.098	11:33:16
Air Blank	0.000	11:33:45

Preventative Maintenance  
Performed -- Timer Reset