

COPY

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murby AGENCY Yavapai Co. S.O.
DATE 2/24/13 TIME 12:36:37
INTOXILYZER SERIAL # 90-01734 LOCATION Seligman

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 160 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 098 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murby NS6/5597

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

02/24/2013

12:36:37

Standard Lot#: 09312100A1

Last Changed By: J. MURRAY #N56

QAS: J. MURRAY #N56

YAVAPAI YCSO

Operator: J. MURRAY #N56

YAVAPAI YCSO

Test g/210L Time

Air Blank 0.000 12:37:12
0.100 Cal Check 0.098 12:37:32
Air Blank 0.000 12:38:02

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
FORM PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

CASE NO. 2013-01734 AGENCY YAVAPAI CO. S.O.
DATE 6/12/2013 TIME 12:41:30
INTOXILYZER SERIAL # 80-01734 LOCATION SELIGMAN

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Empty recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

QA SPECIALIST [Signature]
DATE 7 9 00 PST 09312100A1
INTOXILYZER 6/12/2014

DIAGNOSTIC TESTS
SIGNATURE J. Murray NS6/5597

DPS Form EXHIBIT G-4 (REV 05-01)

- 1. Alcohol-free subject
- 2. Empty recognition logic system functioning
- 3. Proper sample recognition system
- 4. Standard calibration

Instrument is operating properly and accurately.

COMMENTS

OXILYZER 8000
Station: YCSO SELIGMAN
Serial Number: 80-001734
Version: 8105.48
10/20/2013

12:41:30

Standard Lot#: 09312100A1
Changed By: J. MURRAY #N56
J. MURRAY #N56
PAI YCSO
Operator: J. MURRAY #N56
PAI YCSO

Object: SQAP,TEST
Date: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:42:28
Diagnostic Test	Pass	12:43:03
Air Blank	0.000	12:43:30
100 Cal Check	0.098	12:43:51
Air Blank	RFI*	12:44:07
Air Blank	IPS**	12:44:17

RFI Detect
Improper Sample

Sex:
Date:
Time:

15 11

Test

10

10

10

10

10

10

10

Not a Successfully Completed Test Sequence

KILYZER 8000
Location: YCSO SELIGMAN
Serial Number: 80-001734
Version: 8105.48
Date: 10/20/2013

12:44:33

Card Lot#: 09312100A1
Changed By: J. MURRAY #N56
J. MURRAY #N56
YCSO
Operator: J. MURRAY #N56
YCSO

Subject: TEST, SOAP
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:45:48
Diagnostic Test	Pass	12:46:23
Air Blank	0.000	12:46:51
0.100 Cal Check	0.098	12:47:11
Air Blank	0.000	12:47:41
Subject Test	0.000	12:48:01
Air Blank	RFI*	12:48:09
Air Blank	IPS**	12:48:16

RFI Detect
Improper Sample

Not a Successfully Completed Test Sequence

ANALYZER 8000
Location: YCSO SELIGMAN
Serial Number: 80-001734
Firmware Version: 8105.48
Date: 10/20/2013

12:48:29

Card Lot#: 09312100A1
Changed By: J. MURRAY #N56

J. MURRAY #N56
YCSO

Operator: J. MURRAY #N56
YCSO

Subject: SOAP, TEST
DOB: 10/20/1960

Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:49:26
Diagnostic Test	Pass	12:50:01
Air Blank	0.000	12:50:28
4400 Cal Check	0.099	12:50:49
Air Blank	0.000	12:51:18
Subject Test	0.000	12:51:35
Air Blank	0.000	12:52:02
Wait	ABT*	
Air Blank	RFI**	12:52:21

Sequence Aborted
RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray AGENCY Yavapai Co. S.O.
DATE 2/24/2013 TIME 13:01:53
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

- (A) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (A) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 097 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NS6/5597

DPS Form Exh G-2 (Rev 05-01)

QA SPECIALIST J. Murray
DATE 2/24/2013
INTOXILYZER SERIAL # _____

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution _____ AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results _____ AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

02/24/2013

13:01:53

Standard Lot#: 09312100A1

Last Changed By: J. MURRAY #N56

QAS: J. MURRAY #N56

YAVAPAI YCSO

Operator: J. MURRAY #N56

YAVAPAI YCSO

Test	g/210L	Time
Air Blank	0.000	13:02:27
0.100 Cal Check	0.097	13:02:48
Air Blank	0.000	13:03:18

Preventative Maintenance
Performed -- Timer Reset