

COPY

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. MURRAY AGENCY Yavapai County Sheriff's Office
DATE 12/2/2014 TIME 12:00:26
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution 100 AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (+) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(+) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(+) 4. Air blank completed.
(+) 5. Calibration check completed. Test results 0. 098 AC.
(+) 6. Air blank completed.
(+) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NS6/5597

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. Murray AGENCY Yavapai County Sheriff's Office
DATE 12/21/2014 TIME 12:03:59
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

(X) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(X) 1. Clock time check.

(X) 2. Date check.

OPERATIONAL TESTS

(X) 1. Alcohol-free subject test result 0. 000 AC.

(X) 2. Error recognition logic system functioning.

Not a Successfully Completed Test Sequence printed

(X) 3. Proper sample recognition system.

Not a Successfully Completed Test Sequence printed

Deficient sample printed.

(X) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES X NO

COMMENTS

33913100A3

1/1/2016

-900 PSI

SIGNATURE J. Murray NS6/5597

INTOXILYZER 8000
Location: YCSO SELIGMAN
Serial Number: 80-001734
Core Version: 8105.48
12/21/2014

12:03:59

Standard Lot#: 33913100A3
Last Changed By: J. MURRAY5597 #N56

QAS: J. MURRAY5597 #N56
YAVAPAI SHERIFFS

Operator: J. MURRAY5597 #N56
YAVAPAI SHERIFFS

Subject: TEST,TEST,SQAP
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:05:30
Diagnostic Test	Pass	12:06:05
Air Blank	0.000	12:06:32
0.100 Cal Check	0.098	12:06:53
Air Blank	0.000	12:07:22
Subject Test	0.000	12:07:51
Air Blank	0.000	12:08:18
Wait	ABT*	
Air Blank	IPS**	12:08:41

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

12/21/2014

12:08:52

Standard Lot#: 33913100A3

Last Changed By: J. MURRAY5597 #N56

QAS: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Operator: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Subject: SOAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	12:09:33
Air Blank	0.000	12:10:01

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO SELIGMAN
Serial Number: 80-001734
Core Version: 8105.48
12/21/2014

12:11:23

Standard Lot#: 33913100A3
Last Changed By: J. MURRAY5597 #N56

QAS: J. MURRAY5597 #N56
YAVAPAI SHERIFFS

Operator: J. MURRAY5597 #N56
YAVAPAI SHERIFFS

Subject: SOAP,TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:12:24
Diagnostic Test	Pass	12:12:59
Air Blank	0.000	12:13:27
0.100 Cal Check	0.098	12:13:48
Air Blank	0.000	12:14:17
Subject Test	XXX*	12:14:52
Air Blank	0.000	12:15:26

*Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

12/21/2014

12:15:55

Standard Lot#: 33913100A3

Last Changed By: J. MURRAY5597 #N56

OAS: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Operator: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:16:54
Diagnostic Test	Pass	12:17:29
Air Blank	0.000	12:17:56
0.100 Cal Check	0.097	12:18:17
Air Blank	0.000	12:18:46
Subject Test	0.328*	12:21:54
Air Blank	0.000	12:22:26
Air Blank	0.000	12:22:53
Subject Test	XXX**	12:23:16
Air Blank	0.000	12:23:50

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray AGENCY Yavapai County Sheriff's Office
DATE 12/21/2014 TIME 12:24:09
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.096 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray 1156/5597

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

12/21/2014

12:24:09

Standard Lot#: 33913100A3

Last Changed By: J. MURRAY5597 #N56

QAS: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Operator: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Test g/210L Time

Air Blank 0.000 12:24:49
0.100 Cal Check 0.096 12:25:09
Air Blank 0.000 12:25:39

Preventative Maintenance
Performed -- Timer Reset