

COPY

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray AGENCY Yavapai County Sheriff's Office
DATE 3/16/15 TIME 14:53:21
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
Pour a standard alcohol concentration solution OR AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 097 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NS6/5597

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

03/16/2015

14:53:21

Standard Lot#: 33913100A3

Last Changed By: J. MURRAY5597 #N56

QAS: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Operator: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Test	g/210L	Time
Air Blank	0.000	14:54:08
0.100 Cal Check	0.097	14:54:29
Air Blank	0.000	14:54:58

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

F. Murray

AGENCY Yavapai County Sheriff's Office

6/15

TIME 14:56:03

SERIAL # 80-01734

LOCATION Seligman

Display reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock battery check.
- 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.

2. Expiratory recognition logic system functioning.

3. Not a Successfully Completed Test Sequence printed

4. Expiratory sample recognition system.

5. Not a Successfully Completed Test Sequence printed

6. Efficient sample printed.

7. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

1. A. S. P. M. L. S. T.

33913100A3

DATE

XP. 1/1/2016

INSTRUMENT #

850 PSI

1. Display Reads

DIAGNOSTIC TESTS
SIGNATURE

F. Murray NS6/5597

INTOXILYZER 8000
Location: YCSO SELIGMAN
Serial Number: 80-001734
Core Version: 8105.48
03/16/2015

14:56:03

Standard Lot#: 33913100A3
Last Changed By: J. MURRAY5597 #N56

QAS: J. MURRAY5597 #N56
YAVAPAI SHERIFFS

Operator: J. MURRAY5597 #N56
YAVAPAI SHERIFFS

Subject: SOAP,TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:57:00
Diagnostic Test	Pass	14:57:35
Air Blank	0.000	14:58:02
0.100 Cal Check	0.098	14:58:23
Air Blank	0.000	14:58:52
Subject Test	0.000	14:59:17
Air Blank	0.000	14:59:44
Wait	ABT*	
Air Blank	IPS**	15:00:25

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

03/16/2015

15:01:09

Standard Lot#: 33913100A3

Last Changed By: J. MURRAY5597 #N56

OAS: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Operator: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Subject: SOAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	15:01:50
Air Blank	IPS**	15:01:58

*RFI Detect

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Firmware Version: 8105.48

03/16/2015

15:02:17

Standard Lot#: 33913100A3

Last Changed By: J. MURRAY5597 #N56

QAS: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Operator: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:03:41
Diagnostic Test	Pass	15:04:16
Air Blank	0.000	15:04:43
0.100 Cal Check	0.097	15:05:04
Air Blank	0.000	15:05:33
Subject Test	0.483*	15:08:41
Air Blank	0.000	15:09:13
Air Blank	0.000	15:09:41
Subject Test	XXX**	15:10:14
Air Blank	0.000	15:10:48

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

NAME MURPHY AGENCY Yavapai County Sheriff's Office
ID # 5 TIME 15:11:04
INSTRUMENT ID # 80-01734 LOCATION Seligman

(*) 1. Ensure the gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR

Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- (+) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (+) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made, press ENTER.
- (+) 4. Air blank completed.
- (+) 5. Calibration check completed. Test results 0. 097 AC.
- (+) 6. Air blank completed.
- (+) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE MURPHY NSB/5597

DPS Form Exh G-2 (Rev 05-01)

04 SPECIALIST

CAUSE

INTOXILYZER SERIAL

(*) 1. Ensure that

2. Intoxilyzer

3. Go to the

4. Air blank

5. Calibration

6. Air blank

7. Remove

SIGNATURE

DPS Form Exh G-2 (R

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

03/16/2015

15:11:04

Standard Lot#: 33913100A3

Last Changed By: J. MURRAY5597 #N56

QAS: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Operator: J. MURRAY5597 #N56

YAVAPAI SHERIFFS

Test	g/210L	Time
Air Blank	0.000	15:11:46
0.100 Cal Check	0.097	15:12:07
Air Blank	0.000	15:12:37

Preventative Maintenance
Performed -- Timer Reset