

DR 17-028107

COPY

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Scott Blakley Sr. AGENCY Yavapai County SO
DATE 8-6-17 TIME 23:09:59
INTOXILYZER SERIAL # 80-01734 LOCATION Seligman

- (d) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution NA AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (x) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (x) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (x) 4. Air blank completed.
- (x) 5. Calibration check completed. Test results 0. 098 AC.
- (x) 6. Air blank completed.
- (x) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Scott Blakley Sr.

DPS Form Exh G-2 (Rev 05-01) Tank Serial # 22315100A1
Exp 9-5-2017
Press. 300 psi
- Past 31 day Check
- Cal Check out of Tolerance

DR 17-028107

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

08/06/2017

23:01:30

Standard Lot#: 22315100A1

Last Changed By: J. MURRAY5597 #N56

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	23:02:07
0.100 Cal Check	0.000*	23:02:51
Air Blank	0.000	23:03:23

*Calibration Check Out of Tolerance

Selected Wet Test ~~test~~ by mistake

PR 17-028107

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

08/06/2017

23:06:34

Standard Lot#: 22315100A1

Last Changed By: J. MURRAY5597 #N56

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	23:07:12
0.100 Cal Check	0.098	23:07:33
Air Blank	0.000	23:08:02

Preventative Maintenance
Performed -- Timer Reset

DR 17-028107

INTOXILYZER 8000
Location: YCSO SELIGMAN
Serial Number: 80-001734
Core Version: 8105.48
08/06/2017 23:09:59

Standard Lot#: 22315100A1
Last Changed By: J. MURRAY5597 #N56

QAS: S. BLAKLEY SR #L6861
YAVAPAI CSO

Operator: S. BLAKLEY SR #L6861
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	23:10:29
0.100 Cal Check	0.098	23:10:50
Air Blank	0.000	23:11:20

Preventative Maintenance
Performed -- Timer Reset