

DR 18-039205

COPY

EXHIBIT G-2  
 THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)  
 ARIZONA DEPARTMENT OF PUBLIC SAFETY  
 STANDARD QUALITY ASSURANCE PROCEDURES  
 INTOXILYZER MODEL 8000  
 STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Scott Blakley Sr. AGENCY Yavapai County SO  
 DATE 10-25-69 TIME 21:08:42  
 INTOXILYZER SERIAL # 80-01734 LOCATION YCSO Seligman

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
 OR  
 Pour a standard alcohol concentration solution NA AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.098 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

DPS Form Exh G-2 (Rev 05-01)

Tank Serial # 06717100A5  
 Exp 5-5-2019  
 Pres 850 psi

Beyond 31 day by 4 day  
 Due to time off.

DR 18-039205

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

10/25/2018

21:05:15

-----  
Standard Lot#: 0617100A5

Last Changed By: S. BLAKLEY SR #L6861

-----  
QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO

-----  
Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	21:05:45
0.100 Cal Check	0.098	21:06:06
Air Blank	0.000	21:06:36

Preventative Maintenance  
Performed -- Timer Reset

DR, 18-039205

INTOXILYZER 8000

Location: YCSO SELIGMAN

Serial Number: 80-001734

Core Version: 8105.48

10/25/2018

21:08:42

-----  
Standard Lot#: 0617100A5

Last Changed By: S. BLAKLEY SR #L6861  
-----

QAS: S. BLAKLEY SR #L6861

YAVAPAI CSO  
-----

Operator: S. BLAKLEY SR #L6861

YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	21:09:12
0.100 Cal Check	0.098	21:09:32
Air Blank	0.000	21:10:02

Preventative Maintenance  
Performed -- Timer Reset