

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Jacobs AGENCY CVMO
DATE 12/17/14 TIME 1307
INTOXILYZER SERIAL # 80-001736 LOCATION CVMO - Booking

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0009 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  CV23 D. Jacobs

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001736
Core Version: 8105.48
12/17/2014 13:09:30

Standard Lot#: 14014100A1
Last Changed By: S. SOTO #6432

QAS: D. JACOBS #CV23
CAMP VERDE MO

Operator: D. JACOBS #CV23
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	13:10:04
0.100 Cal Check	0.099	13:10:25
Air Blank	0.000	13:10:55

Preventative Maintenance
Performed -- Timer Reset