

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Deborah Hughes AGENCY Camp Verde Marshals Ofc
DATE 1-16-15 TIME 1102
INTOXILYZER SERIAL # 80-001736 LOCATION CVMO - Booking

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 0006 AC:
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 00 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE [Signature] #53

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001736
Core Version: 8105.48
01/16/2015 11:02:32

Standard Lot#: 14014100A1
Last Changed By: S. SOTO #6432

QAS: D. HUGHES #53
CAMP VERDE MO

Operator: D. HUGHES #53
CAMP VERDE MO

Subject: TEST,TEST
DOB: 01/16/1975
Sex: F Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:04:04
Diagnostic Test	Pass	11:04:38
Air Blank	0.000	11:05:05
0.100 Cal Check	0.099	11:05:26
Air Blank	0.000	11:05:56
Subject Test	0.000	11:06:22
Air Blank	0.000	11:06:50
Wait	IPS*	
Air Blank	0.000	11:08:22

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

01/16/2015

11:10:18

Standard Lot#: 14014100A1

Last Changed By: S. SOTO

#6432

QAS: D. HUGHES

#53

CAMP VERDE MO

Operator: D. HUGHES

#53

CAMP VERDE MO

Subject: TEST,TEST

DOB: 01/16/1975

Sex: F

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:11:30
Diagnostic Test	Pass	11:12:04
Air Blank	0.000	11:12:32
0.100 Cal Check	0.099	11:12:53
Air Blank	0.000	11:13:22
Subject Test	0.185*	11:16:29
Air Blank	0.000	11:17:09
Air Blank	0.000	11:17:36
Subject Test	XXX**	11:18:17
Air Blank	RFI***	11:18:30

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

01/16/2015

11:18:54

Standard Lot#: 14014100A1

Last Changed By: S. SOTO

#6432

QAS: D. HUGHES

#53

CAMP VERDE MO

Operator: D. HUGHES

#53

CAMP VERDE MO

Subject: TEST,TEST

DOB: 01/16/1975

Sex: F

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	ABT*	11:19:45
Air Blank	0.000	11:20:13

*Sequence Aborted

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Deborah Hughes AGENCY Camp Verde Marshals Ofc
DATE 1-16-15 TIME 1120
INTOXILYZER SERIAL # 80-001736 LOCATION CVMO-Booking

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.097 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] # 53

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

01/16/2015

11:20:54

Standard Lot#: 14014100A1

Last Changed By: S. SOTO #6432

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Test g/210L Time

Air Blank 0.000 11:21:25
0.100 Cal Check 0.097 11:21:45
Air Blank 0.000 11:22:15

Preventative Maintenance
Performed -- Timer Reset