

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Jacobs AGENCY Camp Verde Marshals office
DATE 1/13/16 TIME 2131
INTOXILYZER SERIAL # 80-001736 LOCATION Cvmo Evidence Room

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 099 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.


SIGNATURE  CV23

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST D. Jacobs AGENCY Camp Verde Marshals office
DATE 1/13/16 TIME 2113
INTOXILYZER SERIAL # 80-001736 LOCATION Cvmo Evidence room

(X) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (X) 1. Clock time check.
(X) 2. Date check.

OPERATIONAL TESTS

- (X) 1. Alcohol-free subject test result 0. 000 AC.
(X) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(X) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(X) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

Dop Toponek present as operator while I conducted QAS

SIGNATURE  CVZ3

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

01/13/2016

21:19:15

Standard Lot#: 14014100A1

Last Changed By: S. SOTO

#6432

QAS: D. JACOBS

#CV23

CAMP VERDE MO

Operator: M. TOPOREK

#CV20

CAMP VERDE MO

Subject: TEST,TEST

DOB: 01/01/1960

Sex: M

Weight: 175

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 21:20:29 |
| Diagnostic Test | Pass | 21:21:04 |
| Air Blank | 0.000 | 21:21:31 |
| 0.100 Cal Check | 0.099 | 21:21:52 |
| Air Blank | 0.000 | 21:22:22 |
| Subject Test | 0.054* | 21:25:29 |
| Air Blank | 0.000 | 21:26:01 |
| Air Blank | 0.000 | 21:26:28 |
| Subject Test | XXX** | 21:27:35 |
| Air Blank | RFI*** | 21:27:43 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

01/13/2016

21:13:33

Standard Lot#: 14014100A1

Last Changed By: S. SOTO

#6432

QAS: D. JACOBS

#CV23

CAMP VERDE MO

Operator: M. TOPOREK

#CV20

CAMP VERDE MO

Subject: TEST,TEST

DOB: 01/01/1960

Sex: M

Weight: 175

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 21:15:10 |
| Diagnostic Test | Pass | 21:15:45 |
| Air Blank | 0.000 | 21:16:13 |
| 0.100 Cal Check | 0.098 | 21:16:34 |
| Air Blank | 0.000 | 21:17:03 |
| Subject Test | 0.000 | 21:17:40 |
| Air Blank | 0.000 | 21:18:08 |
| Wait | ABT* | |
| Air Blank | IPS** | 21:18:26 |

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D Jacobs AGENCY Camp Verde Marshals office
DATE 1/13/16 TIME 2111
INTOXILYZER SERIAL # 80-001736 LOCATION Cvmo Evidence Room

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(X) 4. Air blank completed.
(X) 5. Calibration check completed. Test results 0. 099 AC.
(X) 6. Air blank completed.
(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  CV23

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

01/13/2016

21:11:22

Standard Lot#: 14014100A1

Last Changed By: S. SOTO #6432

QAS: D. JACOBS #CV23

CAMP VERDE MO

Operator: D. JACOBS #CV23

CAMP VERDE MO

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 21:12:01 |
| 0.100 Cal Check | 0.099 | 21:12:22 |
| Air Blank | 0.000 | 21:12:52 |

Preventative Maintenance
Performed -- Timer Reset