

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Jacobs AGENCY Camp Verde Marshals
DATE 4/21/16 TIME 2039
INTOXILYZER SERIAL # 80-001736 LOCATION Cvmo DUI Trailer

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(X) 4. Air blank completed.
(X) 5. Calibration check completed. Test results 0. 100 AC.
(X) 6. Air blank completed.
(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] CV2B

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

04/21/2016

20:39:08

Standard Lot#: 14014100A1

Last Changed By: S. SOTO #6432

QAS: D. JACOBS #CV23

CAMP VERDE MO

Operator: D. JACOBS #CV23

CAMP VERDE MO

Test g/210L Time

Air Blank 0.000 20:50:54
0.100 Cal Check 0.100 20:51:14
Air Blank 0.000 20:51:44

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST D. Jacobs AGENCY Camp Verde Marshals
DATE 4/21/16 TIME 2053
INTOXILYZER SERIAL # 80-001736 LOCATION CVRO DUI Trailer

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE [Signature]

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

04/21/2016

20:53:23

Standard Lot#: 14014100A1

Last Changed By: S. SOTO

#6432

QAS: D. JACOBS

#CV23

CAMP VERDE MO

Operator: D. JACOBS

#CV23

CAMP VERDE MO

Subject: TEST,TEST

DOB: 01/30/1991

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	20:54:29
Diagnostic Test	Pass	20:55:04
Air Blank	0.000	20:55:32
0.100 Cal Check	0.100	20:55:53
Air Blank	0.000	20:56:22
Subject Test	0.000	20:56:55
Air Blank	ABT*	20:57:08
Air Blank	IPS**	20:57:16

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001736
Core Version: 8105.48
04/21/2016

20:58:13

Standard Lot#: 14014100A1
Last Changed By: S. SOTO #6432

QAS: D. JACOBS #CV23
CAMP VERDE MO

Operator: D. JACOBS #CV23
CAMP VERDE MO

Subject: TEST,TEST
DOB: 01/30/1991
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	20:59:16
Diagnostic Test	Pass	20:59:51
Air Blank	0.000	21:00:18
0.100 Cal Check	0.100	21:00:39
Air Blank	0.000	21:01:08
Subject Test	0.058*	21:04:16
Air Blank	0.000	21:04:47
Air Blank	0.000	21:05:14
Subject Test	XXX**	21:06:42
Air Blank	RFI***	21:06:51

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Jacobs AGENCY Camp Verde Marshals
DATE 4/21/16 TIME 2108
INTOXILYZER SERIAL # 80-001736 LOCATION Cyano DUE Trailer

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(X) 4. Air blank completed.
(X) 5. Calibration check completed. Test results 0. 099 AC.
(X) 6. Air blank completed.
(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  CV23

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

04/21/2016

21:08:10

Standard Lot#: 14014100A1

Last Changed By: S. SOTO #6432

QAS: D. JACOBS #CV23

CAMP VERDE MO

Operator: D. JACOBS #CV23

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	21:08:41
0.100 Cal Check	0.099	21:09:02
Air Blank	0.000	21:09:32

Preventative Maintenance
Performed -- Timer Reset