

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Deborah Hughes AGENCY Camp Verde Marshal's Ofc
DATE 5/26/17 TIME 0942
INTOXILYZER SERIAL # 80-001736 LOCATION CVMO-Booking Ofc

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.098 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] #63

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

05/26/2017

09:44:20

Standard Lot#: 09216100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #63

CAMP VERDE MO

Operator: D. HUGHES #63

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	09:45:11
0.100 Cal Check	0.098	09:45:32
Air Blank	0.000	09:46:02

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Deborah Hughes AGENCY Camp Verde Marshal's Ofc
DATE 5/26/17 TIME 0947
INTOXILYZER SERIAL # 80-001734 LOCATION Cumo-Booking Ofc

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC:
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. _____ AC.

Instrument is operating properly and accurately. YES NO _____

COMMENTS

SIGNATURE DJ #63

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001736
Core Version: 8105.48
05/26/2017

09:47:15

Standard Lot#: 09216100A1
Last Changed By: D. HUGHES #53

QAS: D. HUGHES #63
CAMP VERDE MO

Operator: D. HUGHES #63
CAMP VERDE MO

Subject: TEST,TEST
DOB: 05/27/1963
Sex: F Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:48:38
Diagnostic Test	Pass	09:49:12
Air Blank	0.000	09:49:40
0.100 Cal Check	0.098	09:50:00
Air Blank	0.000	09:50:30
Subject Test	0.000	09:50:50
Air Blank	ABT*	09:50:58
Air Blank	IPS**	09:51:07

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001736
Core Version: 8105.48
05/26/2017 09:51:40

Standard Lot#: 09216100A1
Last Changed By: D. HUGHES #53

QAS: D. HUGHES #63
CAMP VERDE MO

Operator: D. HUGHES #63
CAMP VERDE MO

Subject: TEST,TEST
DOB: 05/27/1963
Sex: F Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:52:47
Diagnostic Test	Pass	09:53:21
Air Blank	0.000	09:53:48
0.100 Cal Check	0.098	09:54:09
Air Blank	0.000	09:54:38
Subject Test	RGE*	09:55:09
Air Blank	0.000	09:55:43

*Range Exceeded

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001736
Core Version: 8105.48
05/26/2017 09:56:09

Standard Lot#: 09216100A1
Last Changed By: D. HUGHES #53

QAS: D. HUGHES #63
CAMP VERDE MO

Operator: D. HUGHES #63
CAMP VERDE MO

Subject: TEST,TEST
DOB: 05/27/1963
Sex: F Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:57:12
Diagnostic Test	Pass	09:57:46
Air Blank	0.000	09:58:13
0.100 Cal Check	0.098	09:58:34
Air Blank	0.000	09:59:04
Subject Test	0.029*	10:02:11
Air Blank	RFI**	10:02:29
Air Blank	0.000	10:02:57

*Deficient Sample
**RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Deborah Hughes AGENCY Camp Verde Marshal's Ofc
DATE 5/20/17 TIME 1003
INTOXILYZER SERIAL # 80-001734 LOCATION C/MD - Booking Ofc

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.1120 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] # 103

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001736

Core Version: 8105.48

05/26/2017

10:04:34

Standard Lot#: 09216100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #63

CAMP VERDE MO

Operator: D. HUGHES #63

CAMP VERDE MO

Test g/210L Time

Air Blank 0.000 10:05:08
0.100 Cal Check 0.098 10:05:29
Air Blank 0.000 10:05:59

Preventative Maintenance
Performed -- Timer Reset