

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Sgt. D. Jacobs AGENCY Camp Verde Marshals
DATE 6/24/18 TIME 1213
INTOXILYZER SERIAL # 80-001736 LOCATION CVMO DUI Trailer

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] CV08

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001736
Core Version: 8105.48
06/24/2018 12:13:04

Standard Lot#: 32316100A1
Last Changed By: D. JACOBS #CV08

QAS: D. JACOBS #CV08
CAMP VERDE MO

Operator: D. JACOBS #CV08
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	12:13:34
0.100 Cal Check	0.099	12:13:55
Air Blank	0.000	12:14:25

Preventative Maintenance
Performed -- Timer Reset