

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST BOLING, L #11323 AGENCY YAVAPAI COUNTY SHERIFF'S OFFICE
DATE 07/27/2026 TIME 1825
INTOXILYZER SERIAL # 9000-2811

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0.096 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES X NO

COMMENTS:

SIGNATURE Mike Boling #11323

26-036955

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-002811
Software Version: 9439.01.00
Analytical Report Number: 28110727260166

QAS: SMITH, MCKENNA
QAS Permit #: 30459
Agency: YAVAPAI CSO
Last 31-Day Check: 07/02/2026
Last Annual Maintenance: 10/03/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 07/27/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 18:10
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: BOLING, LUKE
Agency: YAVAPAI CSO

Permit #: 30212

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-403434231
Expiration Date: 10/03/2027
Bottle #: 21
Last Changed By: KAUFMAN, JASON

Permit #: 30402

Test	g/210L	Time
Air Blank	0.000	18:25:39
Diagnostic Test	PASS	18:26:14
Air Blank	0.000	18:26:51
Calibration Chk	0.096	18:27:13
Air Blank	0.000	18:27:53
Subject Test 1	DEF*	18:31:05
Air Blank	0.000	18:31:58
Air Blank	0.000	18:32:38
Subject Test 2	0.000	18:33:45
Air Blank	0.000	18:34:26
Wait		18:38:31
Air Blank	0.000	18:39:08
Subject Test 3	INV**	18:39:27
Air Blank	RFI***	18:39:40

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTSEXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.