

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST D M'Ginnis AGENCY YCSO
DATE 4-25-09 TIME 2049
INTOXILYZER SERIAL # 80-000789 LOCATION Mobile Command (YCSO)
Williamson Valley Rd BUICK Pontiac

DM 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
DM 1. Clock time check.
DM 2. Date check.

OPERATIONAL TESTS
DM 1. Alcohol-free subject test result 0. 00 AC.
DM 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
DM 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
DM 4. Standard Calibration Check standard 0. 10 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES DM NO

COMMENTS All OK

SIGNATURE [Signature] S43/GAS# 33426

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. McGinnis AGENCY YCSO

DATE 4-25-09 TIME 2103

INTOXILYZER SERIAL # 80-000289 LOCATION Mobile Command YCSO Williamson Valley RD

- [Signature]* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution _____ AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. ~~Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
- [Signature]* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- [Signature]* 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- [Signature]* 4. Air blank completed.
- [Signature]* 5. Calibration check completed. Test results 0. 097 AC.
- [Signature]* 6. Air blank completed.
- [Signature]* 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE *[Signature]* 543/QAS# 33426

INSTRUMENT #800
 LOCATION: VCSO-MOSBILE
 SERIAL NUMBER: 80-000289
 09/25/2015 20:53:00

Standard Lot#: 9014011
 Last Charged By: D. REISS #3692
 09/25/2015 20:49:33

Operator: D. MCCANNIS #8086
 YAGAPP, CSC

Operator: D. MCCANNIS #8086
 YAGAPP, CSC

Subject: SOAP, TEST
 DOB: 09/19/1990
 Sex: M Weight: 210

Subject: SOAP, TEST
 DOB: 09/19/1990
 Sex: M Weight: 210

15 Minute Deprivation Period? Yes

15 Minute Deprivation Period? Yes

Test	g/2:0L	Time
Air Blank	0.000	20:54:45
Diagnostic Test	Pass	20:55:14
Air Blank	0.000	20:55:43
0.100 Cal Check	0.096	20:56:00
Air Blank	0.000	20:56:29
Subject Test	0.058*	20:59:37
Air Blank	0.000	21:00:09
Air Blank	0.000	21:00:36
Subject Test	XXX**	21:01:01
Air Blank	IPS***	21:01:11

Test	g/2:0L	Time
Air Blank	0.000	20:51:13
Diagnostic Test	Pass	20:51:41
Air Blank	0.000	20:52:10
0.100 Cal Check	0.099	20:52:27
Air Blank	0.000	20:52:57
Subject Test	0.000	20:53:16
Air Blank	RP*	20:53:24
Air Blank	RF:**	20:53:31

*Deficient Sample
 **Invalid Sample
 Begin new deprivation period
 ***Improper Sample

*Sequence Aborted
 **RF: Detect

Not a Successfully
 Completed
 Test Sequence

Not a Successfully
 Completed
 Test Sequence

[Signature]
 5473/33474

INSTRUMENT #800
 LOCATION: VCSO-MOSBILE
 SERIAL NUMBER: 80-000289
 09/25/2015 20:49:33

Standard Lot#: 9014011
 Last Charged By: D. REISS #3692
 09/25/2015 20:49:33

Operator: D. MCCANNIS #8086
 YAGAPP, CSC

Subject: SOAP, TEST
 DOB: 09/19/1990
 Sex: M Weight: 210

Test	g/2:0L	Time
Air Blank	0.000	20:47:02
Cal Check	0.097	20:47:21
Air Blank	0.000	20:47:50

INTELLIZER 8005
Location: YCSO-MOBILE
Serial Number: 80-00289
04/25/2005

INTELLIZER 8005
Location: YCSO-MOBILE
Serial Number: 80-00289
04/25/2005

Standard Lot#: 9014011
Last Charged By: D. RAISS #3693
04/25/2005

Standard Lot#: 9014011
Last Charged By: D. RAISS #3693
04/25/2005

QAS: S. MCINNIS #8086
YACAPAI CSC

QAS: D. MCINNIS #8086
YACAPAI CSC

Test	9/21/11	Time
Air Blank	0.000	21:02:49
Cal Check	0.097	21:04:08
Air Blank	0.005	21:04:37

Test	9/21/11	Time
Air Blank	0.000	21:02:49
Cal Check	0.097	21:04:08
Air Blank	0.005	21:04:37

[Signature]
Operator's Signature

[Signature]
Operator's Signature