

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST DAN RAISS AGENCY YAVAPAI Co.S.O.
DATE 8-14-10 TIME 1700
INTOXILYZER SERIAL # 80-000289 LOCATION YCSO-MOBILE-CIB

- DR) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- DR) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- DR) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- DR) 4. Air blank completed.
- DR) 5. Calibration check completed. Test results 0. 095 AC.
- DR) 6. Air blank completed.
- DR) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel Q. Raiss # C-10/33026

DPS Form Exh G-2 (Rev 05-01)

UNLOCK INSTRUMENT FOR USE AT DRE TRAINING ALCOHOL
WORKSHOP.

INTOXILYZER 8000
Location: YCSO-MOBILE
Serial Number: 80-000289
Core Version: 8105.48
08/14/2010 07:07:32

Standard Lot#: 712322
Last Changed By: D. RAISS #3693

DAS: D. RAISS #3693
YAUAPAI CSO

Operator: D. RAISS #3693
YAUAPAI CSO

Test	g/210L	Time
Air Blank	0.000	07:08:12
0.100 Cal Check	0.095	07:08:33
Air Blank	0.000	07:09:02

Preventative Maintenance
Performed -- Timer Reset