

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.  
DATE 11/20/2012 TIME 1142  
INTOXILYZER SERIAL # 80-000289 LOCATION YCSO DUI TRAILER

- (  ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (  ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (  ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (  ) 4. Air blank completed.
- (  ) 5. Calibration check completed. Test results 0. 095 AC.
- (  ) 6. Air blank completed.
- (  ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody 7486

INTOXILYZER 8000  
Instrument Initialization  
07:33:29 11/20/2012

INTOXILYZER 8000  
Location: YCSD DUI TRAILER  
Serial Number: 80-000289  
Core Version: 8105.48  
11/20/2012 11:42:10

-----  
Standard Lot#: 13712100A2  
Last Changed By: L. MOODY #7486  
-----

DAS: L. MOODY #7486  
YAUAPAI CSO  
-----

Operator: L. MOODY #7486  
YAUAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	11:42:42
0.100 Cal Check	0.095	11:43:03
Air Blank	0.000	11:43:32

Preventative Maintenance  
Performed -- Timer Reset