

COPY

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.

DATE 2/21/12 TIME 1810

INTOXILYZER SERIAL # 80-000289 LOCATION YCSO DUI TRAILER

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 095 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE LMoody 74886

COPY

INTOXILYZER 8000
Instrument Initialization
17:39:50 12/21/2012

INTOXILYZER 8000
Location: YCSO DUI TRAILER
Serial Number: 80-000289
Core Version: 8105.48
12/21/2012 18:10:38

Standard Lot#: 13712100A2
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAUAPAI CSO

Operator: L. MOODY #7486
YAUAPAI CSO

Test	g/210L	Time
Air Blank	0.000	18:11:07
0.100 Cal Check	0.095	18:11:28
Air Blank	0.000	18:11:58

Preventative Maintenance
Performed -- Timer Reset