

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.  
DATE 6/29/2015 TIME 2104  
INTOXILYZER SERIAL # 80-000239 LOCATION YCSO - BLACK CANYON

- (07) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.  
OR  
~~Pour a standard alcohol concentration solution \_\_\_\_\_ AC. into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C + 0.2°C~~
- (07) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (07) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (07) 4. Air blank completed.
- (07) 5. Calibration check completed. Test results 0. 100 AC.
- (07) 6. Air blank completed.
- (07) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody

INTOXILYZER 8002  
Location: YCSO - BLACK CANYON  
Serial Number: 80-000289  
Core Version: 8105.48  
06/29/2015 21:04:14

-----  
Standard Lot#: 25414100A4  
Last Changed By: L. MOODY #7486  
-----

QAS: L. MOODY #7486  
YAUAPAI CSO  
-----

Operator: L. MOODY #7486  
YAUAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	21:04:46
0.100 Cal Check	0.100	21:05:07
Air Blank	0.000	21:05:36

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.  
DATE 6/29/2015 TIME 2107  
INTOXILYZER SERIAL # 80-000289 LOCATION YCSO - BLACK CANYON

(67) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (67) 1. Clock time check.  
(67) 2. Date check.

OPERATIONAL TESTS

- (67) 1. Alcohol-free subject test result 0. 000 AC.  
(67) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(67) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(67) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES 67 NO \_\_\_\_\_

COMMENTS

SIGNATURE Logan Moody

INTOXILYZER 8000  
Location: YCSO - BLACK CANYON  
Serial Number: 80-000289  
Core Version: 8105.48  
06/29/2015 21:07:39

Standard Lot#: 25414100A4  
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486  
YAUAPAI CSO

Operator: L. MOODY #7486  
YAUAPAI CSO

Subject: SQAP,TEST  
DOB: 01/01/2000  
Sex: M Weight: 190

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:08:37
Diagnostic Test	Pass	21:09:14
Air Blank	0.000	21:09:42
0.100 Cal Check	0.100	21:10:02
Air Blank	0.000	21:10:32
Subject Test	0.000	21:10:56
Air Blank	ABT*	21:11:09
Air Blank	RFI**	21:11:16

\*Sequence Aborted  
\*\*RFI Detect

Not a Successfully  
Completed  
Test Sequence

INTOXILYZER 8000  
Location: YCSO - BLACK CANYON  
Serial Number: 80-000289  
Core Version: 8105.48  
06/29/2015 21:11:35

Standard Lot#: 25414100A4  
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486  
YAUAPAI CSO

Operator: L. MOODY #7486  
YAUAPAI CSO

Subject: SQAP,TEST  
DOB: 01/01/2000  
Sex: M Weight: 190

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:12:25
Diagnostic Test	Pass	21:13:02
Air Blank	0.000	21:13:30
0.100 Cal Check	0.100	21:13:51
Air Blank	0.000	21:14:20
Subject Test	0.056*	21:17:28
Air Blank	0.000	21:17:59
Air Blank	0.000	21:18:26
Subject Test	XXX**	21:18:47
Air Blank	IPS***	21:18:54

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period  
\*\*\*Improper Sample

Not a Successfully  
Completed  
Test Sequence

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.

DATE 6/29/2015 TIME 2119

INTOXILYZER SERIAL # 80-000239 LOCATION YCSO - BLACK CANON

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (M) 4. Air blank completed.
- (M) 5. Calibration check completed. Test results 0.100 AC.
- (M) 6. Air blank completed.
- (M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody 7486

INTOXILYZER 8000  
Location: YCSO - BLACK CANYON  
Serial Number: 00-000285  
Core Version: 8105.48  
06/29/2015 21:19:26

Standard Lot#: 25414100A4  
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486  
YAUAPAI CSO

Operator: L. MOODY #7486  
YAUAPAI CSO

Test	g/210L	Time
Air Blank	0.000	21:19:55
0.100 Cal Check	0.100	21:20:16
Air Blank	0.000	21:20:45

Preventative Maintenance  
Performed -- Timer Reset