

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YCSO

DATE 08/08/2015 TIME 2029

INTOXILYZER SERIAL # 80-000289 LOCATION YCSO-BLACK CANYON

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
~~Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 099 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody 2486

INTOXILYZER 8000
Location: CSI - BLACK CANYON
Serial Number: 60-060265
Core Version: 0105.48
08-08-2015 20:29:46

Standard Lot#: 25414109A4
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
VAUAPAI CSO

Operator: L. MOODY #7486
VAUAPAI CSO

Test	g/210L	Time
Air Blank	0.000	20:30:33
0.100 Cal Check	0.199	20:30:54
Air Blank	0.000	20:31:23

Preventative Maintenance
Performed -- Timer Reset