

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YCSO

DATE 08/29/2015 TIME 2306

INTOXILYZER SERIAL # 80-000289 LOCATION YCSO - BLACK CANYON

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (M) 4. Air blank completed.
- (M) 5. Calibration check completed. Test results 0. 100 AC.
- (M) 6. Air blank completed.
- (M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody

INTOXILYZER 8000
Instrument Initialization
18:50:46 08/29/2015

INTOXILYZER 8000
Location: YCSG - BLACK CANYON
Serial Number: 80-000289
Core Version: 8105.48
08/29/2015 23:06:36

Standard Lot#: 25414100A4
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAUAPAI CSO

Operator: L. MOODY #7486
YAUAPAI CSO

Test	g/210L	Time
Air Blank	0.000	23:07:07
0.100 Cal Check	0.100	23:07:28
Air Blank	0.000	23:07:57

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY GCSD
DATE 8/29/2015 TIME 2309
INTOXILYZER SERIAL # 40-000289 LOCATION GCSD-BLACK CANYON

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

1000 psi

SIGNATURE Logan Moody 7-886

INTOXILYZER 8000
Location: YCSO - BLACK CANYON
Serial Number: 80-000289
Core Version: 8105.48
08/29/2015 23:09:32

Standard Lot#: 25414100A4
Last Changed By: L. MOODY #7486

OAS: L. MOODY #7486
YAUAPAI CSO

Operator: L. MOODY #7486
YAUAPAI CSO

Subject: SQAP, TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:10:22
Diagnostic Test	Pass	23:10:59
Air Blank	0.000	23:11:27
0.100 Ca! Check	0.100	23:11:48
Air Blank	0.000	23:12:17
Subject Test	0.000	23:12:52
Air Blank	ABT*	23:12:58
Air Blank	RFI**	23:13:06

*Sequence Aborted
**RFI Detect

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: YCSO - BLACK CANYON
Serial Number: 80-000289
Core Version: 8105.48
08/29/2015 23:13:35

Standard Lot#: 25414100A4
Last Changed By: L. MOODY #7486

OAS: L. MOODY #7486
YAUAPAI CSO

Operator: L. MOODY #7486
YAUAPAI CSO

Subject: SQAP, TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:14:22
Diagnostic Test	Pass	23:14:59
Air Blank	0.000	23:15:27
0.100 Ca! Check	0.099	23:15:48
Air Blank	0.000	23:16:17
Subject Test	0.095*	23:19:25
Air Blank	0.000	23:19:56
Air Blank	0.000	23:20:23
Subject Test	XXX**	23:20:52
Air Blank	!PS***	23:21:00

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***Improper Sample

Not a Successfully
Completed
Test Sequence