

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOEAN MOODY 07486 AGENCY YAVAPAI CO. S.O.
DATE 10/27/2015 TIME 1955
INTOXILYZER SERIAL # 80-000289 LOCATION 9050- BLACK CANYON

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
~~Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$~~
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 100 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE LOEAN MOODY 07486

RECEIVED

NOV 12 2015

INTOX:LYZER 8000
Location: YCSO - BLACK CANYON
Serial Number: 80-000289
Core Version: 8105.48
10/27/2015 19:55:48

Standard Lot#: 25414100A4
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAUAPA! CSO

Operator: L. MOODY #7486
YAUAPA! CSO

Test	g/210L	Time
Air Blank	0.000	19:56:18
0.100 Cal Check	0.100	19:56:39
Air Blank	0.000	19:57:08

Preventative Maintenance
Performed -- Timer Reset