

18-010390

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Logan Moody #7486 AGENCY Yavapai County Sheriff's Office
DATE 3/23/2018 TIME 1128
INTOXILYZER SERIAL # 80-000289 LOCATION YCSO - Black Canyon

- (M)* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
~~Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
- (M)* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(M) 4. Air blank completed.
(M) 5. Calibration check completed. Test results 0. 097 AC.
(M) 6. Air blank completed.
(M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE *Logan Moody #7486*

18-010390

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Logan Moody #7486 AGENCY Yavapai County Sheriff's Office

DATE 3/23/2018 TIME 1130

INTOXILYZER SERIAL # 80-000289 LOCATION YCSO - Black Canyon

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.

2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.

2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed

3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.

4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

900 psi

SIGNATURE Logan Moody #7486

18-010390

INTOXILYZER 8000

Location: YCSO - BLACK CANYON

Serial Number: 80-000289

Core Version: 8105.48

03/23/2018

11:30:52

Standard Lot#: 23016100A1

Last Changed By: L. MOODY

#7486

QAS: L. MOODY

#7486

YAVAPAI CSO

Operator: L. MOODY

#7486

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 01/01/1990

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:31:39
Diagnostic Test	Pass	11:32:14
Air Blank	0.000	11:32:42
0.100 Cal Check	0.098	11:33:03
Air Blank	0.000	11:33:32
Subject Test	0.000	11:33:53
Air Blank	ABT*	11:34:00
Air Blank	RFI**	11:34:07

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

18-010390

INTOXILYZER 8000

Location: YCSO - BLACK CANYON

Serial Number: 80-000289

Core Version: 8105.48

03/23/2018

11:34:16

Standard Lot#: 23016100A1

Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486

YAVAPAI CSO

Operator: L. MOODY #7486

YAVAPAI CSO

Subject: SQAP, TEST

DOB: 01/01/1990

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:35:02
Diagnostic Test	Pass	11:35:37
Air Blank	0.000	11:36:05
0.100 Cal Check	0.097	11:36:25
Air Blank	0.000	11:36:55
Subject Test	0.100*	11:40:02
Air Blank	0.000	11:40:34
Air Blank	0.000	11:41:01
Subject Test	XXX**	11:41:28
Air Blank	IPS***	11:41:35

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***Improper Sample

Not a Successfully Completed Test Sequence

18-010390

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Logan Moody #7486 AGENCY Yavapai County Sheriff's Office
DATE 3/23/2018 TIME 1141
INTOXILYZER SERIAL # 80-000289 LOCATION YCSO - Black Canyon

- (L) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (L) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (L) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (L) 4. Air blank completed.
- (L) 5. Calibration check completed. Test results 0. 097 AC.
- (L) 6. Air blank completed.
- (L) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody

18-010390

INTOXILYZER 8000
Location: YCSO - BLACK CANYON
Serial Number: 80-000289
Core Version: 8105.48
03/23/2018 11:41:50

Standard Lot#: 23016100A1
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Test g/210L Time

Air Blank 0.000 11:42:19
0.100 Cal Check 0.097 11:42:39
Air Blank 0.000 11:43:09

Preventative Maintenance
Performed -- Timer Reset