

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QAS# 33034

QA SPECIALIST MICHAEL MORRIS #50 AGENCY PRESCOTT VALLEY POLICE DEPT.  
DATE 06/26/09 TIME 1936  
INTOXILYZER SERIAL # 80000293 LOCATION PVPD BOOKING ROOM

- (mm) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (mm) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (mm) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (mm) 4. Air blank completed.
- (mm) 5. Calibration check completed. Test results 0. 100 AC.
- (mm) 6. Air blank completed.
- (mm) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Cpl. M. Morris #50

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

06/26/2009

19:36:46


-----  
Standard Lot#: 80441I

Last Changed By: J. TOBIN #73  
-----

QAS: M. MORRIS #50

PRESCOTT VALLEY PD  
-----

Test	g/210L	Time
Air Blank	0.000	19:37:14
Cal Check	0.100	19:37:33
Air Blank	0.000	19:38:02

  
Operator's Signature