

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Chris Krogedal AGENCY PVPD  
DATE 9/2/09 TIME 0721329  
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD Booking

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 099 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Chris Krogedal #90 QAS #33200

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

09/02/2009

13:30:54

-----  
Standard Lot#: 804402I

Last Changed By: M. MORRIS #50  
-----

QAS: C. KROGEDAL #6535

YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	13:31:21
Cal Check	0.099	13:31:40
Air Blank	0.000	13:32:09

*C. Krogedal #90*

Operator's Signature

QAS #33200

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Chris Krogedal #90 AGENCY PVPD  
DATE 9/2/09 TIME 1343  
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD bookings

(  ) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  ) 1. Clock time check.  
(  ) 2. Date check.

OPERATIONAL TESTS

- (  ) 1. Alcohol-free subject test result 0. 000 AC.  
(  ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(  ) 4. Standard Calibration Check standard 0. 099 AC. Result 0. \_\_\_\_\_ AC.

Instrument is operating properly and accurately. YES  NO \_\_\_\_\_

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Chris Krogedal #90 QAS#33200

INTOXILYZER 8000  
Location: PRESCOTT VALLEY PD  
Serial Number : 80-000293  
09/02/2009 13:38:48

---

Standard Lot#: 804402I  
Last Changed By: M. MORRIS #50

---

QAS: C. KROGEDAL #6535  
YAVAPAI CSO

---

Operator: C. KROGEDAL #6535  
YAVAPAI CSO

---

Subject: SQAP, TEST  
DOB: 12/02/1990  
Sex: M Weight: 160

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:40:37
Diagnostic Test	Pass	13:41:05
Air Blank	0.000	13:41:32
0.100 Cal Check	0.099	13:41:51
Air Blank	0.000	13:42:20
Subject Test	0.000	13:42:50
Air Blank	0.000	13:43:17
Five Minute Wait	ABT*	
Air Blank	0.000	13:43:52

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT VALLEY PD  
Serial Number : 80-000293  
09/02/2009 13:43:59

---

Standard Lot#: 804402I  
Last Changed By: M. MORRIS #50

---

QAS: C. KROGEDAL #6535  
YAVAPAI CSO

---

Operator: C. KROGEDAL #6535  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 01/14/1996  
Sex: M Weight: 160

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	13:44:39
Air Blank	0.000	13:45:08

\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT VALLEY PD  
Serial Number : 80-000293  
09/02/2009 13:45:29

---

Standard Lot#: 804402I  
Last Changed By: M. MORRIS #50

---

QAS: C. KROGEDAL #6535  
YAVAPAI CSO

---

Operator: C. KROGEDAL #6535  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 01/14/1967  
Sex: M Weight: 150

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	13:46:09
Air Blank	0.000	13:46:37

\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT VALLEY PD  
Serial Number : 80-000293  
09/02/2009 13:46:52

---

Standard Lot#: 804402I  
Last Changed By: M. MORRIS #50

---

QAS: C. KROGEDAL #6535  
YAVAPAI CSO

---

Operator: C. KROGEDAL #6535  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 01/14/1955  
Sex: M Weight: 150

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:47:46
Diagnostic Test	Pass	13:48:13
Air Blank	0.000	13:48:40
0.100 Cal Check	0.099	13:48:59
Air Blank	0.000	13:49:28
Subject Test	0.347*	13:52:37
Air Blank	0.000	13:53:09
Air Blank	0.000	13:53:37
Subject Test	XXX**	13:53:57
Air Blank	0.000	13:54:30

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**