

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST C. Krogedal #90 AGENCY PVPD
DATE 1/5/10 TIME 1619
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 096 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE C. Krogedal #90 / QAS #33200

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

01/05/2010

16:23:03


Standard Lot#: 585220

Last Changed By: A. LAWRENCE #61

QAS: C. KROGEDAL #6535

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	16:23:30
Cal Check	0.096	16:23:49
Air Blank	0.000	16:24:18



Operator's Signature

QAS # 33200