

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST C. Krogedal AGENCY PVPD
DATE 2/23/10 TIME 1915
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD Booking Room

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 096 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE C. Krogedal #90/QAS # 33200

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

02/23/2010

19:21:52

Standard Lot#: 585220

Last Changed By: A. LAWRENCE #61

QAS: C. KROGEDAL #90

PRESCOTT VALLEY PD

Test	g/210L	Time
Air Blank	0.000	19:22:19
Cal Check	0.096	19:22:38
Air Blank	0.000	19:23:08


Operator's Signature

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST C. Krogedal AGENCY PVPP
DATE 2/23/10 TIME 1915
INTOXILYZER SERIAL # 80-000293 LOCATION PVPP booking room

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 096 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE

C. Krogedal #90 / QAS # 33200

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number : 80-000293
02/23/2010 19:24:00

Standard Lot#: 585220
Last Changed By: A. LAWRENCE #61

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST
DOB: 01/23/1975
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:25:17
Diagnostic Test	Pass	19:25:45
Air Blank	0.000	19:26:12
0.100 Cal Check	0.096	19:26:31
Air Blank	0.000	19:27:00
Subject Test	0.000	19:27:26
Air Blank	0.000	19:27:53
Five Minute Wait	ABT*	
Air Blank	0.000	19:28:28

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number : 80-000293
02/23/2010 19:29:59

Standard Lot#: 585220
Last Changed By: A. LAWRENCE #61

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST
DOB: 01/23/1975
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	19:30:43
Air Blank	0.000	19:31:11

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number : 80-000293

02/23/2010

19:31:20

Standard Lot#: 585220

Last Changed By: A. LAWRENCE #61

QAS: C. KROGEDAL #90

PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90

PRESCOTT VALLEY PD

Subject: SQAP,TEST

DOB: 01/23/1975

Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	19:31:59
Air Blank	0.000	19:32:28

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number : 80-000293
02/23/2010 19:33:01

Standard Lot#: 585220
Last Changed By: A. LAWRENCE #61

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST
DOB: 01/23/1975
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:34:16
Diagnostic Test	Pass	19:34:43
Air Blank	0.000	19:35:10
0.100 Cal Check	0.090	19:35:29
Air Blank	0.000	19:35:58
Subject Test	0.291*	19:39:07
Air Blank	0.000	19:39:39
Air Blank	0.000	19:40:06
Subject Test	XXX**	19:40:22
Air Blank	0.000	19:40:54

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence