

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST C. Krogedal AGENCY PVPD  
DATE 2/24/10 TIME 1235  
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD Bodily Room

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.  
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.  
() 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
() 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
() 4. Standard Calibration Check standard 0. 097 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE C. Krogedal #90/QAS #33200

INTOXILYZER 8000  
Location: PRESCOTT VALLEY PD  
Serial Number: 80-000293  
Core Version: 8105.48  
02/24/2010

12:36:11

---

Standard Lot#: 585220  
Last Changed By: A. LAWRENCE #61

---

QAS: C. KROGEDAL #90  
PRESCOTT VLYPD

---

Operator: C. KROGEDAL #90  
PRESCOTT VALLEY PD

---

Subject: SQAP,TEST  
DOB: 01/23/1956  
Sex: M Weight: 150

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:37:03
Diagnostic Test	Pass	12:37:37
Air Blank	0.000	12:38:04
0.100 Cal Check	0.097	12:38:25
Air Blank	0.000	12:38:54
Subject Test	0.000	12:40:02
Air Blank	0.000	12:40:29
Wait	ABT*	
Air Blank	0.000	12:41:05

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

02/24/2010

12:41:48

---

Standard Lot#: 585220

Last Changed By: A. LAWRENCE #61

---

QAS: C. KROGEDAL #90

PRESCOTT VLYPD

---

Operator: C. KROGEDAL #90

PRESCOTT VALLEY PD

---

Subject: SQAP,TEST

DOB: 01/23/1967

Sex: M Weight: 150

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	12:42:32
Air Blank	0.000	12:43:00

\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT VALLEY PD  
Serial Number: 80-000293  
Core Version: 8105.48  
02/24/2010 12:43:56

---

Standard Lot#: 585220  
Last Changed By: A. LAWRENCE #61

---

QAS: C. KROGEDAL #90  
PRESCOTT VLYPD

---

Operator: C. KROGEDAL #90  
PRESCOTT VALLEY PD

---

Subject: SQAP,TEST  
DOB: 01/23/1957  
Sex: M Weight: 150

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	12:44:34
Air Blank	0.000	12:45:02

\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT VALLEY PD  
Serial Number: 80-000293  
Core Version: 8105.48  
02/24/2010 12:46:07

---

Standard Lot#: 585220  
Last Changed By: A. LAWRENCE #61

---

QAS: C. KROGEDAL #90  
PRESCOTT VLYPD

---

Operator: C. KROGEDAL #90  
PRESCOTT VALLEY PD

---

Subject: SQAP,TEST  
DOB: 01/23/1957  
Sex: M Weight: 150

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:47:14
Diagnostic Test	Pass	12:47:48
Air Blank	0.000	12:48:15
0.100 Cal Check	0.098	12:48:36
Air Blank	0.000	12:49:05
Subject Test	0.170*	12:52:12
Air Blank	0.000	12:52:44
Air Blank	0.000	12:53:12
Subject Test	0.111**	12:53:44
Air Blank	0.000	12:54:16
Wait		
Air Blank	0.000	12:59:11
Subject Test	0.179**	12:59:50
Air Blank	0.000	13:00:25

\*Deficient Sample

\*\*No .020 agreement

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT VALLEY PD  
Serial Number: 80-000293  
Core Version: 8105.48  
02/24/2010 13:05:53

---

Standard Lot#: 585220  
Last Changed By: A. LAWRENCE #61

---

QAS: C. KROGEDAL #90  
PRESCOTT VLYPD

---

Operator: C. KROGEDAL #90  
PRESCOTT VALLEY PD

---

Subject: SQAP,TEST  
DOB: 01/23/1956  
Sex: M Weight: 150

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:06:42
Diagnostic Test	Pass	13:07:16
Air Blank	0.000	13:07:44
0.100 Cal Check	0.096	13:08:05
Air Blank	0.000	13:08:34
Subject Test	0.011	13:09:10
Air Blank	0.000	13:09:39
Wait		
Air Blank	0.000	13:14:38
Subject Test	XXX*	13:15:47
Air Blank	0.000	13:16:20

\*Invalid Sample  
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST C. Krogedal AGENCY PVPD  
DATE 2/24/10 TIME 1130"  
INTOXILYZER SERIAL # 80-000293 LOCATION YCSO Booking - Prescott

- (  ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (  ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (  ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (  ) 4. Air blank completed.
- (  ) 5. Calibration check completed. Test results 0.095 AC.
- (  ) 6. Air blank completed.
- (  ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE C. Krogedal #90 / QAS# 33200

INTOXILYZER 8000  
Instrument Initialization  
11:17:27 02/24/2010

INTOXILYZER 8000  
Location: PRESCOTT VALLEY PD  
Serial Number: 80-000293  
Core Version: 8105.48  
02/24/2010 11:29:19

-----  
Standard Lot#: 585220  
Last Changed By: A. LAWRENCE #61  
-----

WAS: C. KROGEDAL #90  
PRESCOTT ULYPD  
-----

Operator: C. KROGEDAL #90  
PRESCOTT ULYPD  
-----

Test	g/210L	Time
Air Blank	0.000	11:30:39
0.100 Cal Check	0.095	11:30:59
Air Blank	0.000	11:31:29

Preventative Maintenance  
Performed -- Timer Reset