

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST C. Krogedal #90 AGENCY PVPD

DATE 6/3/10 TIME 1355

INTOXILYZER SERIAL # 80-000293 LOCATION PVPD Booking

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.097 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE C. Krogedal #90 / QAS # 33200

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

06/03/2010

13:56:32

Standard Lot#: 685215

Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90

PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90

PRESCOTT VALLEY PD

Test g/210L Time

Air Blank 0.000 13:57:09
0.100 Cal Check 0.097 13:57:29
Air Blank 0.000 13:57:59

Preventative Maintenance

Performed -- Timer Reset

C Krogedal #90 / QAS # 33200

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST C. Krogedal #90 AGENCY PUPD
DATE 6/3/10 TIME 1420
INTOXILYZER SERIAL # 80-000293 LOCATION PUPD Booking

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 097 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

Dry Gas Standard Changed w/ no issues.

SIGNATURE

C. Krogedal #90 / QAS #33200

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number: 80-000293
Core Version: 8105.48
06/03/2010 14:19:07

Standard Lot#: 685215
Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST
DOB: 01/15/1967
Sex: M Weight: 156

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 14:20:07 |
| Diagnostic Test | Pass | 14:20:41 |
| Air Blank | 0.000 | 14:21:09 |
| 0.100 Cal Check | 0.097 | 14:21:30 |
| Air Blank | 0.000 | 14:21:59 |
| Subject Test | 0.000 | 14:22:20 |
| Air Blank | 0.000 | 14:22:48 |
| Wait | ABT* | |
| Air Blank | 0.000 | 14:24:03 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number: 80-000293
Core Version: 8105.48
06/03/2010 14:24:59

Standard Lot#: 685215
Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST
DOB: 01/15/1967
Sex: M Weight: 156

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | IPS* | 14:25:34 |
| Air Blank | 0.000 | 14:26:02 |

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

06/03/2010

14:26:09

Standard Lot#: 685215

Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90

PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90

PRESCOTT VALLEY PD

Subject: SQAP,TEST

DOB: 01/15/1967

Sex: M Weight: 156

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 14:26:44 |
| Air Blank | 0.000 | 14:27:12 |

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number: 80-000293
Core Version: 8105.48
06/03/2010 14:27:48

Standard Lot#: 685215
Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL *Lot # 685 2#80*
PRESCOTT VALLEY PD *Cyl. 18*

Operator: C. KROGEDAL *.100* #90
PRESCOTT VALLEY PD

Subject: SOAP, TEST
DOB: 01/15/1967
Sex: M Weight: *105.6 28/11*

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 14:28:40 |
| Diagnostic Test | Pass | 14:29:14 |
| Air Blank | 0.000 | 14:29:42 |
| 0.100 Cal Check | 0.097 | 14:30:02 |
| Air Blank | 0.000 | 14:30:32 |
| Subject Test | 0.077* | 14:33:39 |
| Air Blank | 0.000 | 14:34:10 |
| Air Blank | 0.000 | 14:34:38 |
| Subject Test | XXX** | 14:34:54 |
| Air Blank | 0.000 | 14:35:28 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence