

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST C. Krogedal #90 AGENCY PVPD
DATE 12/3/10 TIME 2255
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD Booking

- (C) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (C) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (C) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (C) 4. Air blank completed.
- (C) 5. Calibration check completed. Test results 0. 096 AC.
- (C) 6. Air blank completed.
- (C) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE C. Krogedal #90 / QAS 33200



INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

12/03/2010

23:39:27

Standard Lot#: 779386

Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90

PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90

PRESCOTT VALLEY PD

Test	g/210L	Time
Air Blank	0.000	23:39:56
0.100 Cal Check	0.096	23:40:17
Air Blank	0.000	23:40:47

Preventative Maintenance
Performed -- Timer Reset

17.12.2010
C. KROGEDAL #90
QAS # 33200

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST C. Krogedal AGENCY PVPD
DATE 12/3/10 TIME 2350
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 096 AC. Result 0. _____ AC.

Instrument is operating properly and accurately. YES _____ NO _____

COMMENTS

SIGNATURE _____



INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number: 80-000293
Core Version: 8105.48
12/03/2010 23:45:03

Standard Lot#: 779386
Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST
DOB: 01/15/1975
Sex: M Weight: 157

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:46:04
Diagnostic Test	Pass	23:46:38
Air Blank	0.000	23:47:05
0.100 Cal Check	0.095	23:47:26
Air Blank	0.000	23:47:55
Subject Test	0.000	23:48:49
Air Blank	ABT*	23:49:05
Air Blank	0.000	23:49:33

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number: 80-000293
Core Version: 8105.48
12/03/2010 23:49:58

Standard Lot#: 779386
Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST
DOB: 01/15/1967
Sex: M Weight: 159

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	23:50:32
Air Blank	0.000	23:51:00

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number: 80-000293
Core Version: 8105.48
12/03/2010 23:51:31

Standard Lot#: 779386
Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP, TEST
DOB: 01/15/1967
Sex: M Weight: 159

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	23:52:09
Air Blank	0.000	23:52:37

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number: 80-000293
Core Version: 8105.48
12/03/2010 23:53:35

Standard Lot#: 779386
Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST
DOB: 01/15/1967
Sex: M Weight: 159

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:55:11
Diagnostic Test	Pass	23:55:45
Air Blank	0.000	23:56:12
0.100 Cal Check	0.096	23:56:33
Air Blank	0.000	23:57:02
Subject Test	0.042*	00:00:10
Air Blank	0.000	00:00:41
Air Blank	0.000	00:01:08
Subject Test	0.015	00:02:09
Air Blank	0.000	00:02:40
Wait		
Air Blank	0.000	00:07:36
Subject Test	XXX**	00:08:16
Air Blank	0.000	00:08:48

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence