

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST C. Krogedal #90 AGENCY PVPD  
DATE 12/31/10 TIME 2020  
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD Booking

- (CA) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (CA) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (CA) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (CA) 4. Air blank completed.
- (CA) 5. Calibration check completed. Test results 0. 0.96 AC.
- (CA) 6. Air blank completed.
- (CA) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE C. Krogedal #90 / QAS # 33200

INTOXILYZER 8000  
Location: PRESCOTT VALLEY PD  
Serial Number: 80-000293  
Core Version: 8105.48  
12/31/2010 20:22:20

-----  
Standard Lot#: 779386  
Last Changed By: C. KROGEDAL #90

-----  
QAS: C. KROGEDAL #90  
PRESCOTT VALLEY PD

-----  
Operator: C. KROGEDAL #90  
PRESCOTT VALLEY PD  
-----

Test	g/210L	Time
Air Blank	0.000	20:22:57
0.100 Cal Check	0.096	20:23:18
Air Blank	0.000	20:23:47

Preventative Maintenance  
Performed -- Timer Reset